

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081820\
 Data File : VD066308.D
 Acq On : 19 Aug 2020 07:11
 Operator : VA/SY
 Sample : L3712-04
 Misc : 6.93G/5.00ml/MSVOA D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-26572

Quant Time: Aug 19 09:00:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:49:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.97	168	63375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	113880	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	115870	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	52295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	56514	99.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	199.90%	
35) Dibromofluoromethane	7.91	113	42716	59.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.42%	
50) Toluene-d8	10.33	98	121114	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
62) 4-Bromofluorobenzene	12.63	95	54306	56.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.18%	
Target Compounds						
8) Diethyl Ether	3.71	74	1988	7.028	ug/l	96
16) Acetone	4.16	43	7771	72.436	ug/l #	82
20) Methylene Chloride	4.96	84	25497	40.498	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081820\
Data File : VD066308.D
Acq On : 19 Aug 2020 07:11
Operator : VA/SY
Sample : L3712-04
Misc : 6.93G/5.00ml/MSVOA D/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-26572

Quant Time: Aug 19 09:00:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:49:25 2020
Response via : Initial Calibration

