


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD081820\
 Data File : VD066293.D
 Acq On : 18 Aug 2020 23:03
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080520S.M
 Title : SW846 8260
 Last Update : Wed Aug 05 13:49:25 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.0	31688	PASS
75	95	30	60	52.4	92056	PASS
95	95	100	100	100.0	175616	PASS
96	95	5	9	7.1	12423	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	90.6	159029	PASS
175	174	5	9	7.9	12521	PASS
176	174	95	101	97.2	154560	PASS
177	176	5	9	6.8	10518	PASS