

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082021\
 Data File : VD070200.D
 Acq On : 20 Aug 2021 13:54
 Operator : VA/SY
 Sample : M3418-09RE
 Misc : 3.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-15-0510RE

Quant Time: Aug 23 03:52:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	423719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	804556	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	701251	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	198081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	308286	61.860	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.720%	
35) Dibromofluoromethane	7.914	113	291803	56.138	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 112.280%	
50) Toluene-d8	10.337	98	1006490	49.784	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.560%	
62) 4-Bromofluorobenzene	12.626	95	246435	37.008	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 74.020%	
Target Compounds						Qvalue
16) Acetone	4.161	43	43475	47.355	ug/l	98
17) Carbon Disulfide	4.408	76	19295	1.060	ug/l	98
25) 2-Butanone	7.214	43	13215	10.534	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082021\
Data File : VD070200.D
Acq On : 20 Aug 2021 13:54
Operator : VA/SY
Sample : M3418-09RE
Misc : 3.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-15-0510RE

Quant Time: Aug 23 03:52:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

