

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082021\
 Data File : VD070195.D
 Acq On : 20 Aug 2021 11:31
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 23 03:50:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.572	0.445	22.2	82	0.00
3 P	Chloromethane	0.785	0.614	21.8	77	0.00
4 C	Vinyl Chloride	0.846	0.702	17.0#	80	0.00
5 T	Bromomethane	0.532	0.481	9.6	91	0.00
6 T	Chloroethane	0.541	0.489	9.6	86	0.00
7 T	Trichlorofluoromethane	1.016	0.911	10.3	85	0.00
8 T	Diethyl Ether	0.291	0.271	6.9	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.647	0.586	9.4	86	0.00
10 T	Methyl Iodide	0.591	0.585	1.0	81	0.00
11 T	Tert butyl alcohol	0.033	0.029	12.1	86	0.00
12 CM	1,1-Dichloroethene	0.594	0.549	7.6#	84	0.00
13 T	Acrolein	0.037	0.035	5.4	82	0.00
14 T	Allyl chloride	0.886	0.840	5.2	82	0.00
15 T	Acrylonitrile	0.135	0.131	3.0	84	0.00
16 T	Acetone	0.108	0.128	-18.5	105	0.00
17 T	Carbon Disulfide	2.148	1.860	13.4	79	0.00
18 T	Methyl Acetate	0.323	0.299	7.4	81	0.00
19 T	Methyl tert-butyl Ether	1.233	1.221	1.0	83	0.00
20 T	Methylene Chloride	0.968	0.724	25.2#	87	0.00
21 T	trans-1,2-Dichloroethene	0.690	0.651	5.7	83	0.00
22 T	Diisopropyl ether	1.864	1.860	0.2	82	0.00
23 T	Vinyl Acetate	0.845	0.873	-3.3	82	0.00
24 P	1,1-Dichloroethane	1.275	1.190	6.7	84	0.00
25 T	2-Butanone	0.148	0.154	-4.1	88	0.00
26 T	2,2-Dichloropropane	0.993	0.973	2.0	87	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.707	3.5	84	0.00
28 T	Bromochloromethane	0.464	0.430	7.3	88	0.00
29 T	Tetrahydrofuran	0.097	0.097	0.0	83	0.00
30 C	Chloroform	1.254	1.175	6.3#	84	0.00
31 T	Cyclohexane	1.164	1.071	8.0	84	0.00
32 T	1,1,1-Trichloroethane	1.048	0.995	5.1	85	0.00
33 S	1,2-Dichloroethane-d4	0.588	0.564	4.1	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.323	0.326	-0.9	86	0.00
36 T	1,1-Dichloropropene	0.538	0.529	1.7	84	0.00
37 T	Ethyl Acetate	0.195	0.203	-4.1	83	0.00
38 T	Carbon Tetrachloride	0.488	0.487	0.2	86	0.00
39 T	Methylcyclohexane	0.626	0.632	-1.0	84	0.00
40 TM	Benzene	1.560	1.525	2.2	83	0.00
41 T	Methacrylonitrile	0.106	0.110	-3.8	90	0.00
42 TM	1,2-Dichloroethane	0.400	0.386	3.5	83	0.00
43 T	Isopropyl Acetate	0.365	0.354	3.0	81	0.00
44 TM	Trichloroethene	0.377	0.366	2.9	85	0.00
45 C	1,2-Dichloropropane	0.405	0.400	1.2#	84	0.00
46 T	Dibromomethane	0.197	0.194	1.5	85	0.00
47 T	Bromodichloromethane	0.509	0.504	1.0	85	0.00
48 T	Methyl methacrylate	0.174	0.180	-3.4	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082021\
 Data File : VD070195.D
 Acq On : 20 Aug 2021 11:31
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 23 03:50:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	85	0.00
50 S	Toluene-d8	1.256	1.289	-2.6	86	0.00
51 T	4-Methyl-2-Pentanone	0.190	0.197	-3.7	83	0.00
52 CM	Toluene	0.950	0.951	-0.1#	83	0.00
53 T	t-1,3-Dichloropropene	0.451	0.456	-1.1	84	0.00
54 T	cis-1,3-Dichloropropene	0.547	0.557	-1.8	87	0.00
55 T	1,1,2-Trichloroethane	0.279	0.273	2.2	83	0.00
56 T	Ethyl methacrylate	0.317	0.336	-6.0	83	0.00
57 T	1,3-Dichloropropane	0.483	0.475	1.7	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.167	0.184	-10.2	83	0.00
59 T	2-Hexanone	0.127	0.140	-10.2	86	0.00
60 T	Dibromochloromethane	0.319	0.313	1.9	85	0.00
61 T	1,2-Dibromoethane	0.252	0.253	-0.4	84	0.00
62 S	4-Bromofluorobenzene	0.414	0.425	-2.7	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.323	0.308	4.6	83	0.00
65 PM	Chlorobenzene	1.045	1.027	1.7	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.366	1.3	83	0.00
67 C	Ethyl Benzene	1.847	1.910	-3.4#	83	0.00
68 T	m/p-Xylenes	0.716	0.739	-3.2	84	0.00
69 T	o-Xylene	0.646	0.676	-4.6	84	0.00
70 T	Styrene	1.131	1.191	-5.3	83	0.00
71 P	Bromoform	0.181	0.182	-0.6	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.646	3.846	-5.5	85	0.00
74 T	N-amyl acetate	0.786	0.810	-3.1	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.728	0.729	-0.1	86	0.00
76 T	1,2,3-Trichloropropane	0.507	0.496	2.2	87	0.00
77 T	Bromobenzene	0.809	0.810	-0.1	83	0.00
78 T	n-propylbenzene	4.702	4.994	-6.2	84	0.00
79 T	2-Chlorotoluene	2.705	2.755	-1.8	84	0.00
80 T	1,3,5-Trimethylbenzene	3.104	3.242	-4.4	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.195	0.213	-9.2	84	0.00
82 T	4-Chlorotoluene	2.833	2.868	-1.2	83	0.00
83 T	tert-Butylbenzene	2.523	2.636	-4.5	84	0.00
84 T	1,2,4-Trimethylbenzene	3.079	3.221	-4.6	83	0.00
85 T	sec-Butylbenzene	4.042	4.241	-4.9	85	0.00
86 T	p-Isopropyltoluene	3.266	3.448	-5.6	83	0.00
87 T	1,3-Dichlorobenzene	1.719	1.696	1.3	83	0.00
88 T	1,4-Dichlorobenzene	1.703	1.684	1.1	84	0.00
89 T	n-Butylbenzene	3.205	3.399	-6.1	84	0.00
90 T	Hexachloroethane	0.689	0.688	0.1	86	0.00
91 T	1,2-Dichlorobenzene	1.473	1.463	0.7	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.109	0.099	9.2	79	0.00
93 T	1,2,4-Trichlorobenzene	0.819	0.822	-0.4	82	0.00
94 T	Hexachlorobutadiene	0.484	0.485	-0.2	84	0.00
95 T	Naphthalene	1.491	1.539	-3.2	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082021\
Data File : VD070195.D
Acq On : 20 Aug 2021 11:31
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 23 03:50:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.715	0.713	0.3	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6