

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082422\
 Data File : VD074228.D
 Acq On : 24 Aug 2022 17:34
 Operator : VA/SY
 Sample : N4261-01
 Misc : 5.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C5

Quant Time: Aug 25 05:03:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 23 13:12:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	189349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	343601	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	299541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	106658	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	109952	56.827	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.660%
35) Dibromofluoromethane	7.902	113	108094	47.032	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.060%
50) Toluene-d8	10.326	98	355873	48.111	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.220%
62) 4-Bromofluorobenzene	12.614	95	103776	35.513	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	71.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082422\
Data File : VD074228.D
Acq On : 24 Aug 2022 17:34
Operator : VA/SY
Sample : N4261-01
Misc : 5.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C5

Quant Time: Aug 25 05:03:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082322S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 23 13:12:17 2022
Response via : Initial Calibration

