

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082725\
 Data File : VD080502.D
 Acq On : 27 Aug 2025 12:49
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Aug 28 02:42:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082725S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 28 02:42:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	103097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	480197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	607946	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	324163	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	119590	102.815	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	205.640%#
35) Dibromofluoromethane	7.806	113	206634	71.811	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	143.620%
50) Toluene-d8	10.270	98	710017	61.839	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	123.680%
62) 4-Bromofluorobenzene	12.570	95	310354	87.164	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	174.320%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082725\
Data File : VD080502.D
Acq On : 27 Aug 2025 12:49
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Aug 28 02:42:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082725S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 28 02:42:08 2025
Response via : Initial Calibration

