



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD082725\
Data File : VD080495.D
Acq On : 27 Aug 2025 08:09
Operator : RP/MD
Sample : BFB
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D082725S.M
Title : SW846 8260
Last Update : Thu Aug 28 03:02:37 2025



AutoFind: Scans 1835, 1836, 1837; Background Corrected with Scan 1826

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	40205	PASS
75	95	30	60	50.7	114531	PASS
95	95	100	100	100.0	226048	PASS
96	95	5	9	6.8	15334	PASS
173	174	0.00	2	0.9	1554	PASS
174	95	50	100	79.0	178581	PASS
175	174	5	9	7.7	13705	PASS
176	174	95	101	96.2	171755	PASS
177	176	5	9	7.0	11966	PASS