

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083024\
 Data File : VD079753.D
 Acq On : 30 Aug 2024 13:23
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Aug 31 04:42:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	126810	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	262084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	99839	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	93918	54.318	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.640%
35) Dibromofluoromethane	7.804	113	89103	52.305	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.600%
50) Toluene-d8	10.269	98	365229	49.990	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.980%
62) 4-Bromofluorobenzene	12.569	95	93925	45.945	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	91.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083024\
Data File : VD079753.D
Acq On : 30 Aug 2024 13:23
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Aug 31 04:42:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

