

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083024\
 Data File : VD079754.D
 Acq On : 30 Aug 2024 14:43
 Operator : RP/MD
 Sample : VSTDICV050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD083024

Quant Time: Aug 31 04:42:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.600	0.559	6.8	104	0.00
3 P	Chloromethane	1.251	1.157	7.5	102	0.00
4 C	Vinyl Chloride	1.507	1.367	9.3#	96	0.00
5 T	Bromomethane	1.043	0.948	9.1	102	0.00
6 T	Chloroethane	1.022	0.925	9.5	97	0.00
7 T	Trichlorofluoromethane	1.137	1.050	7.7	99	0.00
8 T	Diethyl Ether	0.361	0.324	10.2	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.637	0.599	6.0	101	0.00
10 T	Methyl Iodide	0.734	0.667	9.1	88	0.00
11 T	Tert butyl alcohol	0.034	0.028	17.6	85	0.00
12 CM	1,1-Dichloroethene	0.669	0.647	3.3#	101	0.00
13 T	Acrolein	0.066	0.065	1.5	113	0.00
14 T	Allyl chloride	0.985	0.990	-0.5	103	0.00
15 T	Acrylonitrile	0.163	0.156	4.3	100	0.00
16 T	Acetone	0.141	0.159	-12.8	111	0.00
17 T	Carbon Disulfide	2.445	2.301	5.9	99	0.00
18 T	Methyl Acetate	0.406	0.330	18.7	88	0.00
19 T	Methyl tert-butyl Ether	1.463	1.423	2.7	98	0.00
20 T	Methylene Chloride	0.990	0.799	19.3	97	0.00
21 T	trans-1,2-Dichloroethene	0.742	0.700	5.7	98	0.00
22 T	Diisopropyl ether	2.073	2.073	0.0	98	0.00
23 T	Vinyl Acetate	1.175	1.246	-6.0	106	0.00
24 P	1,1-Dichloroethane	1.356	1.273	6.1	99	0.00
25 T	2-Butanone	0.189	0.189	0.0	103	0.00
26 T	2,2-Dichloropropane	1.055	1.019	3.4	100	0.00
27 T	cis-1,2-Dichloroethene	0.851	0.805	5.4	98	0.00
28 T	Bromochloromethane	0.603	0.608	-0.8	109	0.00
29 T	Tetrahydrofuran	0.117	0.113	3.4	97	0.00
30 C	Chloroform	1.346	1.239	7.9#	96	0.00
31 T	Cyclohexane	1.140	1.094	4.0	98	0.00
32 T	1,1,1-Trichloroethane	1.121	1.048	6.5	96	0.00
33 S	1,2-Dichloroethane-d4	0.682	0.676	0.9	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.325	0.330	-1.5	101	0.00
36 T	1,1-Dichloropropene	0.482	0.487	-1.0	98	0.00
37 T	Ethyl Acetate	0.213	0.214	-0.5	101	0.00
38 T	Carbon Tetrachloride	0.514	0.485	5.6	96	0.00
39 T	Methylcyclohexane	0.587	0.604	-2.9	100	0.00
40 TM	Benzene	1.606	1.540	4.1	96	0.00
41 T	Methacrylonitrile	0.122	0.120	1.6	92	0.00
42 TM	1,2-Dichloroethane	0.394	0.365	7.4	94	0.00
43 T	Isopropyl Acetate	0.397	0.401	-1.0	102	0.00
44 TM	Trichloroethene	0.343	0.324	5.5	99	0.00
45 C	1,2-Dichloropropane	0.388	0.371	4.4#	96	0.00
46 T	Dibromomethane	0.205	0.193	5.9	98	0.00
47 T	Bromodichloromethane	0.512	0.471	8.0	95	0.00
48 T	Methyl methacrylate	0.192	0.191	0.5	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083024\
 Data File : VD079754.D
 Acq On : 30 Aug 2024 14:43
 Operator : RP/MD
 Sample : VSTDICV050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ICVVD083024

Quant Time: Aug 31 04:42:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	79	0.00
50 S	Toluene-d8	1.394	1.456	-4.4	99	0.00
51 T	4-Methyl-2-Pentanone	0.207	0.209	-1.0	100	0.00
52 CM	Toluene	0.972	0.958	1.4#	96	0.00
53 T	t-1,3-Dichloropropene	0.483	0.460	4.8	96	0.00
54 T	cis-1,3-Dichloropropene	0.575	0.556	3.3	96	0.00
55 T	1,1,2-Trichloroethane	0.268	0.248	7.5	95	0.00
56 T	Ethyl methacrylate	0.336	0.341	-1.5	102	0.00
57 T	1,3-Dichloropropane	0.484	0.451	6.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.152	0.159	-4.6	110	0.00
59 T	2-Hexanone	0.144	0.151	-4.9	101	0.00
60 T	Dibromochloromethane	0.330	0.300	9.1	93	0.00
61 T	1,2-Dibromoethane	0.235	0.222	5.5	96	0.00
62 S	4-Bromofluorobenzene	0.390	0.411	-5.4	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.294	0.281	4.4	97	0.00
65 PM	Chlorobenzene	1.109	1.045	5.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.336	7.7	93	0.00
67 C	Ethyl Benzene	1.956	1.956	0.0	96	0.00
68 T	m/p-Xylenes	0.751	0.754	-0.4	95	0.00
69 T	o-Xylene	0.700	0.687	1.9	95	0.00
70 T	Styrene	1.221	1.223	-0.2	94	0.00
71 P	Bromoform	0.183	0.170	7.1	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.559	3.708	-4.2	97	0.00
74 T	N-amyl acetate	0.846	0.919	-8.6	105	0.00
75 P	1,1,2,2-Tetrachloroethane	0.683	0.648	5.1	94	0.00
76 T	1,2,3-Trichloropropane	0.427	0.414	3.0	79	0.00
77 T	Bromobenzene	0.792	0.754	4.8	91	0.00
78 T	n-propylbenzene	4.397	4.650	-5.8	97	0.00
79 T	2-Chlorotoluene	2.450	2.486	-1.5	95	0.00
80 T	1,3,5-Trimethylbenzene	2.958	3.099	-4.8	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.256	0.254	0.8	97	0.00
82 T	4-Chlorotoluene	2.552	2.662	-4.3	97	0.00
83 T	tert-Butylbenzene	2.458	2.576	-4.8	96	0.00
84 T	1,2,4-Trimethylbenzene	3.052	3.155	-3.4	94	0.00
85 T	sec-Butylbenzene	3.759	4.097	-9.0	99	0.00
86 T	p-Isopropyltoluene	3.217	3.461	-7.6	98	0.00
87 T	1,3-Dichlorobenzene	1.700	1.657	2.5	95	0.00
88 T	1,4-Dichlorobenzene	1.700	1.602	5.8	93	0.00
89 T	n-Butylbenzene	3.011	3.303	-9.7	100	0.00
90 T	Hexachloroethane	0.636	0.609	4.2	92	0.00
91 T	1,2-Dichlorobenzene	1.456	1.388	4.7	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.100	0.089	11.0	87	0.00
93 T	1,2,4-Trichlorobenzene	0.800	0.816	-2.0	99	0.00
94 T	Hexachlorobutadiene	0.385	0.399	-3.6	101	0.00
95 T	Naphthalene	1.623	1.695	-4.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083024\
Data File : VD079754.D
Acq On : 30 Aug 2024 14:43
Operator : RP/MD
Sample : VSTDICV050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ICVVD083024

Quant Time: Aug 31 04:42:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.699	0.710	-1.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5