

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091224\
 Data File : VD079826.D
 Acq On : 12 Sep 2024 13:44
 Operator : RP/MD
 Sample : P3900-02
 Misc : 5.02G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VHBLK001

Quant Time: Sep 13 00:17:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 13 00:16:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	153357	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	148857	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	57714	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	151091	22.657	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.640%
7) Chloroethane-d5	2.799	69	127654	24.179	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.720%
11) 1,1-Dichloroethene-d2	3.916	65	28146	23.247	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	93.000%
21) 2-Butanone-d5	6.987	46	22688	36.619	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.240%
24) Chloroform-d	7.569	84	112556	22.328	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.320%
26) 1,2-Dichloroethane-d4	8.228	65	62184	23.590	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.360%
32) Benzene-d6	8.199	84	254716	24.237	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.960%
36) 1,2-Dichloropropane-d6	9.210	67	77053	23.939	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	95.760%
41) Toluene-d8	10.269	98	227796	24.633	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	98.520%
43) trans-1,3-Dichloroprop...	10.522	79	31287	22.610	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.440%
47) 2-Hexanone-d5	10.875	63	19245	36.679	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.360%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	47976	20.284	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.120%
66) 1,2-Dichlorobenzene-d4	13.810	152	56220	25.830	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091224\
Data File : VD079826.D
Acq On : 12 Sep 2024 13:44
Operator : RP/MD
Sample : P3900-02
Misc : 5.02G/10ml/MSVOA_D/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VHBLK001

Quant Time: Sep 13 00:17:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 13 00:16:14 2024
Response via : Initial Calibration

