

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091724\
 Data File : VD079852.D
 Acq On : 17 Sep 2024 15:40
 Operator : RP/MD
 Sample : P4083-01
 Misc : 5.27G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
ClientSampleId :
 A4ES1

Quant Time: Sep 18 04:51:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 18 04:50:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	118898	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	73365	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	14442	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	85372	16.513	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	66.040%		
7) Chloroethane-d5	2.805	69	80128	19.575	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	78.320%		
11) 1,1-Dichloroethene-d2	3.917	65	14530	15.479	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	61.920%		
21) 2-Butanone-d5	6.993	46	14436	30.052	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	60.100%		
24) Chloroform-d	7.563	84	73593	18.830	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	75.320%		
26) 1,2-Dichloroethane-d4	8.234	65	43410	21.240	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	84.960%		
32) Benzene-d6	8.199	84	143011	27.611	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	110.440%		
36) 1,2-Dichloropropane-d6	9.210	67	50209	31.651	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	126.600%#		
41) Toluene-d8	10.269	98	95504	20.954	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	83.800%		
43) trans-1,3-Dichloroprop...	10.528	79	14985	21.972	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	87.880%		
47) 2-Hexanone-d5	10.875	63	10298	39.823	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	79.640%		
56) 1,1,2,2-Tetrachloroeth...	12.646	84	23508	20.167	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	80.680%		
66) 1,2-Dichlorobenzene-d4	13.810	152	9599	17.625	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	70.480%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091724\
 Data File : VD079852.D
 Acq On : 17 Sep 2024 15:40
 Operator : RP/MD
 Sample : P4083-01
 Misc : 5.27G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A4ES1

Quant Time: Sep 18 04:51:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 18 04:50:50 2024
 Response via : Initial Calibration

