

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092022\
 Data File : VD074372.D
 Acq On : 20 Sep 2022 18:16
 Operator : VA/SY
 Sample : N4668-05
 Misc : 6.41G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A2-9-18

Quant Time: Sep 21 02:25:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	103221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	186192	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	142369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	43856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	50520	50.249	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.500%
35) Dibromofluoromethane	7.805	113	52910	43.147	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.300%
50) Toluene-d8	10.269	98	187974	44.425	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.860%
62) 4-Bromofluorobenzene	12.569	95	40033	25.068	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	50.140%
Target Compounds						
16) Acetone	4.028	43	3574	15.779	ug/l	97
17) Carbon Disulfide	4.264	76	7219	2.003	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092022\
Data File : VD074372.D
Acq On : 20 Sep 2022 18:16
Operator : VA/SY
Sample : N4668-05
Misc : 6.41G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A2-9-18

Quant Time: Sep 21 02:25:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

