

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092022\
 Data File : VD074377.D
 Acq On : 20 Sep 2022 20:38
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 21 02:26:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.461	0.399	13.4	72	0.00
3 P	Chloromethane	0.438	0.620	-41.6#	111	0.00
4 C	Vinyl Chloride	0.430	0.380	11.6#	67	0.00
5 T	Bromomethane	0.266	0.114	57.1#	36#	0.00
6 T	Chloroethane	0.253	0.245	3.2	75	0.00
7 T	Trichlorofluoromethane	0.838	0.789	5.8	73	0.00
8 T	Diethyl Ether	0.267	0.284	-6.4	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.527	0.538	-2.1	81	0.00
10 T	Methyl Iodide	0.651	0.401	38.4#	44#	0.00
11 T	Tert butyl alcohol	0.033	0.063	-90.9#	165#	-0.02
12 CM	1,1-Dichloroethene	0.536	0.534	0.4#	78	0.00
13 T	Acrolein	0.035	0.019	45.7#	45#	-0.01
14 T	Allyl chloride	0.776	0.793	-2.2	78	0.00
15 T	Acrylonitrile	0.118	0.132	-11.9	86	0.00
16 T	Acetone	0.110	0.087	20.9	63	0.00
17 T	Carbon Disulfide	1.746	1.460	16.4	64	0.00
18 T	Methyl Acetate	0.326	0.323	0.9	86	0.00
19 T	Methyl tert-butyl Ether	1.148	1.316	-14.6	89	0.00
20 T	Methylene Chloride	0.932	0.776	16.7	87	0.00
21 T	trans-1,2-Dichloroethene	0.594	0.610	-2.7	80	0.00
22 T	Diisopropyl ether	1.634	1.861	-13.9	86	0.00
23 T	Vinyl Acetate	0.748	0.907	-21.3	87	0.00
24 P	1,1-Dichloroethane	1.052	1.162	-10.5	85	0.00
25 T	2-Butanone	0.163	0.158	3.1	78	0.00
26 T	2,2-Dichloropropane	0.868	0.908	-4.6	85	0.00
27 T	cis-1,2-Dichloroethene	0.658	0.710	-7.9	85	0.00
28 T	Bromochloromethane	0.406	0.424	-4.4	78	0.00
29 T	Tetrahydrofuran	0.094	0.103	-9.6	83	0.00
30 C	Chloroform	1.034	1.145	-10.7#	87	0.00
31 T	Cyclohexane	1.021	0.952	6.8	75	0.00
32 T	1,1,1-Trichloroethane	0.913	0.980	-7.3	85	0.00
33 S	1,2-Dichloroethane-d4	0.487	0.511	-4.9	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.329	0.327	0.6	86	0.00
36 T	1,1-Dichloropropene	0.503	0.508	-1.0	80	0.00
37 T	Ethyl Acetate	0.201	0.223	-10.9	86	0.00
38 T	Carbon Tetrachloride	0.445	0.416	6.5	72	0.00
39 T	Methylcyclohexane	0.608	0.593	2.5	76	0.00
40 TM	Benzene	1.409	1.518	-7.7	85	0.00
41 T	Methacrylonitrile	0.109	0.106	2.8	75	0.00
42 TM	1,2-Dichloroethane	0.360	0.379	-5.3	84	0.00
43 T	Isopropyl Acetate	0.372	0.422	-13.4	89	0.00
44 TM	Trichloroethene	0.371	0.396	-6.7	84	0.00
45 C	1,2-Dichloropropane	0.354	0.397	-12.1#	89	0.00
46 T	Dibromomethane	0.182	0.204	-12.1	86	0.00
47 T	Bromodichloromethane	0.449	0.503	-12.0	86	0.00
48 T	Methyl methacrylate	0.180	0.201	-11.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092022\
 Data File : VD074377.D
 Acq On : 20 Sep 2022 20:38
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Sep 21 02:26:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	83	0.00
50 S	Toluene-d8	1.136	1.289	-13.5	83	0.00
51 T	4-Methyl-2-Pentanone	0.195	0.224	-14.9	88	0.00
52 CM	Toluene	0.868	0.956	-10.1#	85	0.00
53 T	t-1,3-Dichloropropene	0.401	0.477	-19.0	90	0.00
54 T	cis-1,3-Dichloropropene	0.518	0.586	-13.1	87	0.00
55 T	1,1,2-Trichloroethane	0.240	0.272	-13.3	88	0.00
56 T	Ethyl methacrylate	0.315	0.360	-14.3	85	0.00
57 T	1,3-Dichloropropane	0.434	0.484	-11.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.080	0.047	41.3#	45#	0.00
59 T	2-Hexanone	0.137	0.149	-8.8	85	0.00
60 T	Dibromochloromethane	0.288	0.317	-10.1	85	0.00
61 T	1,2-Dibromoethane	0.231	0.242	-4.8	82	0.00
62 S	4-Bromofluorobenzene	0.441	0.390	11.6	85	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	84	0.00
64 T	Tetrachloroethene	0.339	0.356	-5.0	87	0.00
65 PM	Chlorobenzene	1.030	1.091	-5.9	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.380	-7.3	88	0.00
67 C	Ethyl Benzene	1.890	2.009	-6.3#	86	0.00
68 T	m/p-Xylenes	0.724	0.764	-5.5	86	0.00
69 T	o-Xylene	0.662	0.736	-11.2	89	0.00
70 T	Styrene	1.127	1.232	-9.3	88	0.00
71 P	Bromoform	0.190	0.190	0.0	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.846	4.072	-5.9	88	0.00
74 T	N-amyl acetate	0.850	0.871	-2.5	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.668	0.698	-4.5	87	0.00
76 T	1,2,3-Trichloropropane	0.463	0.481	-3.9	91	0.00
77 T	Bromobenzene	0.840	0.878	-4.5	89	0.00
78 T	n-propylbenzene	4.770	4.997	-4.8	88	0.00
79 T	2-Chlorotoluene	2.516	2.630	-4.5	88	0.00
80 T	1,3,5-Trimethylbenzene	3.144	3.265	-3.8	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.185	0.192	-3.8	89	0.00
82 T	4-Chlorotoluene	2.609	2.681	-2.8	86	0.00
83 T	tert-Butylbenzene	2.788	2.924	-4.9	88	0.00
84 T	1,2,4-Trimethylbenzene	3.094	3.264	-5.5	89	0.00
85 T	sec-Butylbenzene	4.138	4.383	-5.9	87	0.00
86 T	p-Isopropyltoluene	3.458	3.621	-4.7	88	0.00
87 T	1,3-Dichlorobenzene	1.727	1.779	-3.0	88	0.00
88 T	1,4-Dichlorobenzene	1.688	1.736	-2.8	89	0.00
89 T	n-Butylbenzene	3.387	3.522	-4.0	86	0.00
90 T	Hexachloroethane	0.653	0.549	15.9	73	0.00
91 T	1,2-Dichlorobenzene	1.456	1.518	-4.3	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.099	0.098	1.0	86	0.00
93 T	1,2,4-Trichlorobenzene	0.982	1.016	-3.5	88	0.00
94 T	Hexachlorobutadiene	0.558	0.586	-5.0	86	0.00
95 T	Naphthalene	1.821	1.928	-5.9	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092022\
Data File : VD074377.D
Acq On : 20 Sep 2022 20:38
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 21 02:26:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.844	0.890	-5.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6