

Data Path : W:\HPCHEM1\MSVOA D\DATA\VD092115\
 Data File : VD046799.D
 Acq On : 21 Sep 2015 11:41
 Operator : FY/SY
 Sample : VD0921SBL01
 Misc : 5.00gm/5mL/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0921SBL01

Quant Time: Sep 22 01:32:05 2015
 Quant Method : W:\HPCHEM1\MSVOA_D\METHOD\82D091815S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 18 15:47:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.29	168	184801	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	7.42	114	247354	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.56	117	232908	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	13.90	152	116381	50.00	ug/l	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.70	65	108559	56.50	ug/l	-0.01
Spiked Amount			Recovery	=	113.00%	
35) Dibromofluoromethane	6.20	113	103738	53.60	ug/l	-0.02
Spiked Amount			Recovery	=	107.20%	
50) Toluene-d8	9.61	98	327772	53.69	ug/l	-0.01
Spiked Amount			Recovery	=	107.38%	
62) 4-Bromofluorobenzene	12.84	95	126454	51.44	ug/l	-0.01
Spiked Amount			Recovery	=	102.88%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
21) Methylene Chloride	3.49	84	2839	1.61	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA D\DATA\VD092115\
Data File : VD046799.D
Acq On : 21 Sep 2015 11:41
Operator : FY/SY
Sample : VD0921SBL01
Misc : 5.00qm/5mL/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0921SBL01

Quant Time: Sep 22 01:32:05 2015
Quant Method : W:\HPCHEM1\MSVOA_D\METHOD\82D091815S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 18 15:47:18 2015
Response via : Initial Calibration

