

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092122\
 Data File : VD074385.D
 Acq On : 21 Sep 2022 13:21
 Operator : VA/SY
 Sample : N4641-13RE
 Misc : 5.95G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D22021-18RE

Quant Time: Sep 22 05:12:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	110603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	193178	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	140744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	36852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	49377	45.834	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	91.660%		
35) Dibromofluoromethane	7.804	113	51182	40.228	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	80.460%		
50) Toluene-d8	10.269	98	189547	43.177	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	86.360%		
62) 4-Bromofluorobenzene	12.569	95	35838	20.744	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	41.480%		
Target Compounds						
16) Acetone	4.028	43	3522	14.511	ug/l	92
17) Carbon Disulfide	4.269	76	18259	4.728	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092122\
Data File : VD074385.D
Acq On : 21 Sep 2022 13:21
Operator : VA/SY
Sample : N4641-13RE
Misc : 5.95G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D22021-18RE

Quant Time: Sep 22 05:12:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

