

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092122\
 Data File : VD074388.D
 Acq On : 21 Sep 2022 14:46
 Operator : VA/SY
 Sample : N4641-09
 Misc : 6.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 D22002-17.5

Quant Time: Sep 22 05:12:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 17 04:16:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	114576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	208908	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	181200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	69537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	48977	43.887	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.780%
35) Dibromofluoromethane	7.805	113	50571	36.755	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	73.520%
50) Toluene-d8	10.269	98	195656	41.213	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.420%
62) 4-Bromofluorobenzene	12.569	95	53443	31.051	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	62.100%
Target Compounds						
16) Acetone	4.022	43	4393	17.473	ug/l	89
19) Methyl tert-butyl Ether	5.316	73	10911	4.147	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092122\
Data File : VD074388.D
Acq On : 21 Sep 2022 14:46
Operator : VA/SY
Sample : N4641-09
Misc : 6.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
D22002-17.5

Quant Time: Sep 22 05:12:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D091622S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 17 04:16:54 2022
Response via : Initial Calibration

