

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092524\
 Data File : VD079893.D
 Acq On : 25 Sep 2024 16:34
 Operator : RP/MD
 Sample : P4162-08
 Misc : 5.02G/10ml/MSVOA_D/SOIL/B
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A4ES7

Quant Time: Sep 26 01:57:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 26 01:54:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	180627	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	124183	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	27709	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	130971	16.675	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	66.720%
7) Chloroethane-d5	2.805	69	122177	19.647	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	3.916	65	21672	15.197	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.800%
21) 2-Butanone-d5	6.987	46	14213	19.477	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	38.960%
24) Chloroform-d	7.569	84	103786	17.480	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	69.920%
26) 1,2-Dichloroethane-d4	8.228	65	57726	18.593	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	74.360%
32) Benzene-d6	8.198	84	209747	23.924	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.680%
36) 1,2-Dichloropropane-d6	9.210	67	70179	26.136	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.560%
41) Toluene-d8	10.263	98	150580	19.518	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.080%
43) trans-1,3-Dichloroprop...	10.522	79	18200	15.765	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.080%
47) 2-Hexanone-d5	10.875	63	12045	27.518	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	55.040%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	34681	17.577	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	70.320%
66) 1,2-Dichlorobenzene-d4	13.810	152	18272	17.486	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	69.960%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD092524\
Data File : VD079893.D
Acq On : 25 Sep 2024 16:34
Operator : RP/MD
Sample : P4162-08
Misc : 5.02G/10ml/MSVOA_D/SOIL/B
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A4ES7

Quant Time: Sep 26 01:57:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 26 01:54:46 2024
Response via : Initial Calibration

