

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093024\
 Data File : VD079941.D
 Acq On : 30 Sep 2024 17:15
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Oct 01 05:52:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D093024S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 01 05:47:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	135978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	240294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	218245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	93769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	81246	48.229	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.460%
35) Dibromofluoromethane	7.805	113	75778	44.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.700%
50) Toluene-d8	10.269	98	274061	42.368	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	84.740%
62) 4-Bromofluorobenzene	12.575	95	80428	39.312	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.620%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD093024\
Data File : VD079941.D
Acq On : 30 Sep 2024 17:15
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Oct 01 05:52:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D093024S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 01 05:47:39 2024
Response via : Initial Calibration

