

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101124\
 Data File : VD079961.D
 Acq On : 11 Oct 2024 10:29
 Operator : RP/MD
 Sample : VSTDIC005
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Oct 14 02:48:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 14 02:41:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	146547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	246539	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	224295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	115588	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	9502	5.700	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	11.400%#
35) Dibromofluoromethane	7.805	113	9768	5.489	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	10.980%#
50) Toluene-d8	10.269	98	30268	4.773	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	9.540%#
62) 4-Bromofluorobenzene	12.569	95	10645	4.809	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	9.620%#

Target Compounds				Qvalue		
2) Dichlorodifluoromethane	1.934	85	8830	6.130	ug/l	84
3) Chloromethane	2.152	50	8951	6.051	ug/l	97
4) Vinyl Chloride	2.287	62	10566	5.744	ug/l	95
5) Bromomethane	2.693	94	10327	6.526	ug/l	98
6) Chloroethane	2.834	64	7485	5.961	ug/l	92
7) Trichlorofluoromethane	3.175	101	16789	5.826	ug/l	96
8) Diethyl Ether	3.599	74	4321	5.459	ug/l	89
9) 1,1,2-Trichlorotrifluo...	3.969	101	11125	6.286	ug/l	93
10) Methyl Iodide	4.169	142	8137	4.688	ug/l	96
11) Tert butyl alcohol	5.058	59	2829	27.619	ug/l #	94
12) 1,1-Dichloroethene	3.940	96	8690	5.519	ug/l	90
13) Acrolein	3.787	56	1886	17.251	ug/l #	1
14) Allyl chloride	4.569	41	12293	5.256	ug/l	91
15) Acrylonitrile	5.252	53	10468	27.676	ug/l	96
16) Acetone	4.028	43	12205	28.246	ug/l	97
17) Carbon Disulfide	4.275	76	23706	5.401	ug/l #	94
18) Methyl Acetate	4.564	43	5341	5.758	ug/l	96
19) Methyl tert-butyl Ether	5.305	73	19121	5.258	ug/l #	88
20) Methylene Chloride	4.811	84	14934	4.488	ug/l	97
21) trans-1,2-Dichloroethene	5.305	96	9017	5.333	ug/l	85
22) Diisopropyl ether	6.228	45	26544	5.233	ug/l #	90
23) Vinyl Acetate	6.158	43	71004	24.915	ug/l	100
24) 1,1-Dichloroethane	6.116	63	19217	5.751	ug/l #	92
25) 2-Butanone	7.087	43	15338	29.151	ug/l	97
26) 2,2-Dichloropropane	7.075	77	17110	5.918	ug/l	96
27) cis-1,2-Dichloroethene	7.075	96	11436	5.439	ug/l	92
28) Bromochloromethane	7.422	49	8822	6.185	ug/l	99
29) Tetrahydrofuran	7.452	42	7451	26.478	ug/l	98
30) Chloroform	7.593	83	20683	5.843	ug/l	99
31) Cyclohexane	7.881	56	15026	5.932	ug/l #	73
32) 1,1,1-Trichloroethane	7.793	97	16637	5.674	ug/l	92
36) 1,1-Dichloropropene	8.016	75	12775	5.357	ug/l	98
37) Ethyl Acetate	7.175	43	6299	5.456	ug/l #	95
38) Carbon Tetrachloride	7.987	117	14773	5.574	ug/l	97
39) Methylcyclohexane	9.281	83	13346	4.987	ug/l #	91
40) Benzene	8.252	78	40517	5.372	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101124\
 Data File : VD079961.D
 Acq On : 11 Oct 2024 10:29
 Operator : RP/MD
 Sample : VSTDIC005
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Oct 14 02:48:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 14 02:41:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.393	41	3202	5.281	ug/l #	91
42) 1,2-Dichloroethane	8.322	62	12015	5.793	ug/l	87
43) Isopropyl Acetate	8.363	43	12059	5.669	ug/l #	86
44) Trichloroethene	9.028	130	10086	5.406	ug/l	82
45) 1,2-Dichloropropane	9.304	63	10481	5.480	ug/l	91
46) Dibromomethane	9.399	93	5817	5.513	ug/l	98
47) Bromodichloromethane	9.587	83	15769	5.635	ug/l #	96
48) Methyl methacrylate	9.381	41	4788	4.997	ug/l	95
49) 1,4-Dioxane	9.393	88	1177	101.509	ug/l #	84
51) 4-Methyl-2-Pentanone	10.157	43	27688	25.442	ug/l	95
52) Toluene	10.334	92	22510	4.875	ug/l	95
53) t-1,3-Dichloropropene	10.551	75	12736	5.197	ug/l	92
54) cis-1,3-Dichloropropene	10.016	75	14929	5.163	ug/l	92
55) 1,1,2-Trichloroethane	10.728	97	8022	5.412	ug/l	92
56) Ethyl methacrylate	10.598	69	7751	4.476	ug/l #	88
57) 1,3-Dichloropropane	10.881	76	12688	5.304	ug/l	96
58) 2-Chloroethyl Vinyl ether	9.869	63	13845	21.607	ug/l	96
59) 2-Hexanone	10.922	43	20164	25.017	ug/l	99
60) Dibromochloromethane	11.075	129	10435	5.418	ug/l	99
61) 1,2-Dibromoethane	11.175	107	7591	5.565	ug/l	95
64) Tetrachloroethene	10.810	164	8582	5.425	ug/l #	85
65) Chlorobenzene	11.604	112	28724	5.503	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.681	131	10175	5.426	ug/l	94
67) Ethyl Benzene	11.681	91	43370	4.912	ug/l	99
68) m/p-Xylenes	11.793	106	32330	9.542	ug/l	97
69) o-Xylene	12.122	106	15299	4.880	ug/l	94
70) Styrene	12.134	104	26200	4.697	ug/l	97
71) Bromoform	12.298	173	6457	5.725	ug/l #	97
73) Isopropylbenzene	12.422	105	40318	5.127	ug/l	98
74) N-amyl acetate	12.228	43	9895	5.203	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.669	83	9187	5.762	ug/l	97
76) 1,2,3-Trichloropropane	12.722	75	11840	9.650	ug/l #	100
77) Bromobenzene	12.698	156	11615	5.710	ug/l	94
78) n-propylbenzene	12.763	91	49389	5.052	ug/l	98
79) 2-Chlorotoluene	12.845	91	29511	5.196	ug/l	96
80) 1,3,5-Trimethylbenzene	12.904	105	33333	5.030	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.469	75	3081	5.392	ug/l	91
82) 4-Chlorotoluene	12.951	91	31487	5.160	ug/l	95
83) tert-Butylbenzene	13.163	119	28485	5.015	ug/l	97
84) 1,2,4-Trimethylbenzene	13.210	105	31814	4.781	ug/l	95
85) sec-Butylbenzene	13.340	105	42231	4.838	ug/l	98
86) p-Isopropyltoluene	13.457	119	35260	4.842	ug/l	99
87) 1,3-Dichlorobenzene	13.457	146	22886	5.499	ug/l	99
88) 1,4-Dichlorobenzene	13.534	146	24101	5.796	ug/l	96
89) n-Butylbenzene	13.787	91	36729	5.214	ug/l	95
90) Hexachloroethane	14.051	117	9316	5.787	ug/l	99
91) 1,2-Dichlorobenzene	13.828	146	19930	5.482	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.434	75	1621	6.089	ug/l	88
93) 1,2,4-Trichlorobenzene	15.098	180	11867	5.208	ug/l	99
94) Hexachlorobutadiene	15.198	225	6649	5.299	ug/l	95
95) Naphthalene	15.328	128	20324	5.002	ug/l #	93
96) 1,2,3-Trichlorobenzene	15.516	180	11252	5.572	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101124\
Data File : VD079961.D
Acq On : 11 Oct 2024 10:29
Operator : RP/MD
Sample : VSTDICC005
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :


Quant Time: Oct 14 02:48:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 14 02:41:36 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 14 02:48:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 14 02:41:36 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.875 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion:168 Resp: 146547

Ion Ratio Lower Upper

168 100

99 54.9 46.0 69.0



#2

Dichlorodifluoromethane

Concen: 6.130 ug/l

RT: 1.934 min Scan# 59

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29


Tgt Ion: 85 Resp: 8830

Ion Ratio Lower Upper

85 100

87 41.0 16.1 48.2









#3
Chloromethane
Concen: 6.051 ug/l
RT: 2.152 min Scan# 90
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 50 Resp: 8951
Ion Ratio Lower Upper
50 100
52 37.9 28.8 43.2



#4
Vinyl Chloride
Concen: 5.744 ug/l
RT: 2.287 min Scan# 119
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 62 Resp: 10566
Ion Ratio Lower Upper
62 100
64 30.4 26.5 39.7









#5
Bromomethane
Concen: 6.526 ug/l
RT: 2.693 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 94 Resp: 10327
Ion Ratio Lower Upper
94 100
96 93.0 76.2 114.2



#6
Chloroethane
Concen: 5.961 ug/l
RT: 2.834 min Scan# 212
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 64 Resp: 7485
Ion Ratio Lower Upper
64 100
66 29.8 27.6 41.4









#7
Trichlorofluoromethane
Concen: 5.826 ug/l
RT: 3.175 min Scan# 21
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 16789
Ion Ratio Lower Upper
101 100
103 69.5 53.0 79.6



#8
Diethyl Ether
Concen: 5.459 ug/l
RT: 3.599 min Scan# 342
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 74 Resp: 4321
Ion Ratio Lower Upper
74 100
45 106.0 47.8 143.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 6.286 ug/l
RT: 3.969 min Scan# 405
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 11125
Ion Ratio Lower Upper
101 100
85 41.4 35.3 52.9
151 67.6 60.3 90.5



#10
Methyl Iodide
Concen: 4.688 ug/l
RT: 4.169 min Scan# 439
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:142 Resp: 8137
Ion Ratio Lower Upper
142 100
127 41.4 35.3 52.9
141 13.6 12.2 18.2





#11

Tert butyl alcohol

Concen: 27.619 ug/l

RT: 5.058 min Scan# 59

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 59 Resp: 2829

Ion Ratio Lower Upper

59 100

57 8.0 4.8 7.2#



#12

1,1-Dichloroethene

Concen: 5.519 ug/l

RT: 3.940 min Scan# 400

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion: 96 Resp: 8690


Ion Ratio Lower Upper

96 100

61 140.9 122.6 184.0

98 67.3 47.1 70.7









#13
Acrolein
Concen: 17.251 ug/l
RT: 3.787 min Scan# 31
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 56 Resp: 1886
Ion Ratio Lower Upper
56 100
55 155.4 58.2 87.2#



#14
Allyl chloride
Concen: 5.256 ug/l
RT: 4.569 min Scan# 507
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion: 41 Resp: 12293
Ion Ratio Lower Upper
41 100
39 75.7 53.3 79.9
76 43.2 32.0 48.0








#15
Acrylonitrile
Concen: 27.676 ug/l
RT: 5.252 min Scan# 61
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:	53	Resp:	10468
Ion	Ratio	Lower	Upper
53	100		
52	87.9	66.2	99.4
51	38.3	30.3	45.5



#16
Acetone
Concen: 28.246 ug/l
RT: 4.028 min Scan# 415
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion:	43	Resp:	12205
Ion	Ratio	Lower	Upper
43	100		
58	33.5	25.4	38.2








#17
Carbon Disulfide
Concen: 5.401 ug/l
RT: 4.275 min Scan# 41
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 76 Resp: 23706
Ion Ratio Lower Upper
76 100
78 7.3 7.5 11.3#



#18
Methyl Acetate
Concen: 5.758 ug/l
RT: 4.564 min Scan# 506
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 43 Resp: 5341
Ion Ratio Lower Upper
43 100
74 27.6 20.5 30.7









#19
Methyl tert-butyl Ether
Concen: 5.258 ug/l
RT: 5.305 min Scan# 61
Delta R.T. -0.018 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 73 Resp: 19121
Ion Ratio Lower Upper
73 100
57 17.4 18.8 28.2#



#20
Methylene Chloride
Concen: 4.488 ug/l
RT: 4.811 min Scan# 548
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion: 84 Resp: 14934
Ion Ratio Lower Upper
84 100
49 113.6 86.5 129.7
51 32.4 26.9 40.3
86 58.9 47.2 70.8








#21
trans-1,2-Dichloroethene
Concen: 5.333 ug/l
RT: 5.305 min Scan# 611
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 96	Resp: 9017
Ion Ratio	Lower Upper
96	100
61	114.4 109.4 164.0
98	57.4 49.8 74.8



#22
Diisopropyl ether
Concen: 5.233 ug/l
RT: 6.228 min Scan# 789
Delta R.T. 0.012 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 45	Resp: 26544
Ion Ratio	Lower Upper
45	100
43	43.0 42.6 63.8
87	28.5 22.7 34.1
59	15.1 8.5 12.7#









#23
 Vinyl Acetate
 Concen: 24.915 ug/l
 RT: 6.158 min Scan# 71
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 43 Resp: 71004
 Ion Ratio Lower Upper
 43 100
 86 11.7 9.4 14.0



#24
 1,1-Dichloroethane
 Concen: 5.751 ug/l
 RT: 6.116 min Scan# 770
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Tgt Ion: 63 Resp: 19217
 Ion Ratio Lower Upper
 63 100
 98 5.0 3.5 10.6
 100 8.3 2.1 6.5#





#25

2-Butanone

Concen: 29.151 ug/l

RT: 7.087 min Scan# 91

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 43 Resp: 15338

Ion Ratio Lower Upper

43 100

72 29.3 22.3 33.5



#26

2,2-Dichloropropane

Concen: 5.918 ug/l

RT: 7.075 min Scan# 933

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29

Tgt Ion: 77 Resp: 17110

Ion Ratio Lower Upper

77 100

97 25.8 11.9 35.9









#27
 cis-1,2-Dichloroethene
 Concen: 5.439 ug/l
 RT: 7.075 min Scan# 91
 Delta R.T. -0.006 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 96 Resp: 11436
 Ion Ratio Lower Upper
 96 100
 61 148.7 0.0 278.2
 98 69.8 0.0 127.8



#28
 Bromochloromethane
 Concen: 6.185 ug/l
 RT: 7.422 min Scan# 992
 Delta R.T. -0.006 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Tgt Ion: 49 Resp: 8822
 Ion Ratio Lower Upper
 49 100
 129 0.7 0.0 2.2
 130 82.0 64.6 97.0









#29
Tetrahydrofuran
Concen: 26.478 ug/l
RT: 7.452 min Scan# 996
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 42 Resp: 7451
Ion Ratio Lower Upper
42 100
72 43.5 36.7 55.1
71 43.3 34.4 51.6



#30
Chloroform
Concen: 5.843 ug/l
RT: 7.593 min Scan# 1021
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion: 83 Resp: 20683
Ion Ratio Lower Upper
83 100
85 62.5 50.8 76.2








#31
Cyclohexane
Concen: 5.932 ug/l
RT: 7.881 min Scan# 1070
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 56 Resp: 15026
Ion Ratio Lower Upper
56 100
69 54.5 27.3 40.9#
84 110.0 70.2 105.4#



#32
1,1,1-Trichloroethane
Concen: 5.674 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 97 Resp: 16637
Ion Ratio Lower Upper
97 100
99 70.5 50.7 76.1
61 47.4 34.2 51.2









#33
 1,2-Dichloroethane-d4
 Concen: 5.700 ug/l
 RT: 8.228 min Scan# 1129
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 65 Resp: 9502
 Ion Ratio Lower Upper
 65 100
 67 52.5 0.0 109.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.781 min Scan# 1223
 Delta R.T. 0.006 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Tgt Ion: 114 Resp: 246539
 Ion Ratio Lower Upper
 114 100
 63 19.4 0.0 38.8
 88 14.4 0.0 30.6





#35

Dibromofluoromethane

Concen: 5.489 ug/l

RT: 7.805 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VD079961.D





Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :

Tgt Ion:	113	Resp:	9768
Ion	Ratio	Lower	Upper
113	100		
111	97.7	82.5	123.7
192	15.8	16.6	25.0



#36

1,1-Dichloropropene

Concen: 5.357 ug/l




RT: 8.016 min Scan# 1093


Delta R.T. 0.012 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29


Tgt Ion:	75	Resp:	12775
Ion	Ratio	Lower	Upper
75	100		
110	34.8	18.4	55.0
77	32.3	25.6	38.4








#37
Ethyl Acetate
Concen: 5.456 ug/l
RT: 7.175 min Scan# 91
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 43 Resp: 6299
Ion Ratio Lower Upper
43 100
61 16.9 11.0 16.4#
70 12.7 9.4 14.2



#38
Carbon Tetrachloride
Concen: 5.574 ug/l
RT: 7.987 min Scan# 1088
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 117 Resp: 14773
Ion Ratio Lower Upper
117 100
119 88.4 73.4 110.0
121 30.2 24.5 36.7





#39

Methylcyclohexane

Concen: 4.987 ug/l

RT: 9.281 min Scan# 11308

Delta R.T. 0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion: 83 Resp: 13346

Ion Ratio Lower Upper

83 100

55 73.0 62.0 93.0

98 37.4 38.7 58.1#



#40

Benzene

Concen: 5.372 ug/l

RT: 8.252 min Scan# 1133

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29


Tgt Ion: 78 Resp: 40517

Ion Ratio Lower Upper

78 100

77 25.7 20.2 30.4









#41
Methacrylonitrile
Concen: 5.281 ug/l
RT: 7.393 min Scan# 91
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 41	Resp: 3202
Ion Ratio	Lower Upper
41 100	
39 44.4	47.8 71.8#
67 79.1	64.9 97.3
52 30.9	27.5 41.3



#42
1,2-Dichloroethane
Concen: 5.793 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 62	Resp: 12015
Ion Ratio	Lower Upper
62 100	
98 5.2	0.0 19.8









#43
Isopropyl Acetate
Concen: 5.669 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 12059
Ion Ratio Lower Upper
43 100
61 21.7 26.2 39.2#
87 14.4 11.8 17.8



#44
Trichloroethene
Concen: 5.406 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:130 Resp: 10086
Ion Ratio Lower Upper
130 100
95 79.9 0.0 194.6





#45

1,2-Dichloropropane

Concen: 5.480 ug/l

RT: 9.304 min Scan# 11

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 63 Resp: 10481

Ion Ratio Lower Upper

63 100

65 27.9 26.2 39.4



#46

Dibromomethane

Concen: 5.513 ug/l

RT: 9.399 min Scan# 1328

Delta R.T. 0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion: 93 Resp: 5817


Ion Ratio Lower Upper

93 100

95 85.1 67.9 101.9

174 97.3 80.5 120.7









#47
Bromodichloromethane
Concen: 5.635 ug/l
RT: 9.587 min Scan# 1360
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 15769
Ion Ratio Lower Upper
83 100
85 65.7 50.6 75.8
127 10.5 6.1 9.1#



#48
Methyl methacrylate
Concen: 4.997 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 41 Resp: 4788
Ion Ratio Lower Upper
41 100
69 89.8 74.0 111.0
39 58.4 41.9 62.9





#49

1,4-Dioxane

Concen: 101.509 ug/l

RT: 9.393 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion: 88 Resp: 1177

Ion Ratio Lower Upper

88 100

43 56.4 29.1 43.7#

58 81.6 59.9 89.9



#50

Toluene-d8

Concen: 4.773 ug/l

RT: 10.269 min Scan# 1476

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29


Tgt Ion: 98 Resp: 30268

Ion Ratio Lower Upper

98 100

100 62.9 50.2 75.4









#51
4-Methyl-2-Pentanone
Concen: 25.442 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 27688
Ion Ratio Lower Upper
43 100
58 37.6 32.8 49.2



#52
Toluene
Concen: 4.875 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 92 Resp: 22510
Ion Ratio Lower Upper
92 100
91 181.7 139.3 208.9









#53
t-1,3-Dichloropropene
Concen: 5.197 ug/l
RT: 10.551 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 12736
Ion Ratio Lower Upper
75 100
77 34.5 24.1 36.1



#54
cis-1,3-Dichloropropene
Concen: 5.163 ug/l
RT: 10.016 min Scan# 1433
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion: 75 Resp: 14929
Ion Ratio Lower Upper
75 100
77 35.0 25.0 37.4
39 47.0 33.2 49.8








#55
 1,1,2-Trichloroethane
 Concen: 5.412 ug/l
 RT: 10.728 min Scan# 1554
 Delta R.T. -0.006 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Instrument :
 MSVOA_D
 ClientSampleId :







Tgt Ion:	97	Resp:	8022
Ion	Ratio	Lower	Upper
97	100		
83	79.2	69.2	103.8
85	63.0	43.0	64.6
99	57.0	47.4	71.2



#56
 Ethyl methacrylate
 Concen: 4.476 ug/l
 RT: 10.598 min Scan# 1532
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Tgt Ion:	69	Resp:	7751
Ion	Ratio	Lower	Upper
69	100		
41	72.4	50.4	75.6
39	39.4	26.0	39.0









#57
1,3-Dichloropropane
Concen: 5.304 ug/l
RT: 10.881 min Scan# 1580
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 76 Resp: 12688
Ion Ratio Lower Upper
76 100
78 33.9 25.4 38.2



#58
2-Chloroethyl Vinyl ether
Concen: 21.607 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 63 Resp: 13845
Ion Ratio Lower Upper
63 100
106 29.0 21.8 32.6









#59
2-Hexanone
Concen: 25.017 ug/l
RT: 10.922 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 20164
Ion Ratio Lower Upper
43 100
58 55.0 27.8 83.4



#60
Dibromochloromethane
Concen: 5.418 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:129 Resp: 10435
Ion Ratio Lower Upper
129 100
127 76.0 38.4 115.2





#61

1,2-Dibromoethane

Concen: 5.565 ug/l

RT: 11.175 min Scan# 1630

Delta R.T. -0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion:107 Resp: 7591

Ion Ratio Lower Upper

107 100

109 87.4 73.9 110.9



#62

4-Bromofluorobenzene

Concen: 4.809 ug/l

RT: 12.569 min Scan# 1867

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion: 95 Resp: 10645


Ion Ratio Lower Upper

95 100

174 83.3 0.0 173.4

176 83.5 0.0 166.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:117 Resp: 224295
Ion Ratio Lower Upper
117 100
82 54.6 45.8 68.6
119 33.5 26.2 39.2



#64
Tetrachloroethene
Concen: 5.425 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:164 Resp: 8582
Ion Ratio Lower Upper
164 100
166 116.6 97.5 146.3
129 107.8 70.6 106.0#
131 108.8 68.7 103.1#





#65

Chlorobenzene

Concen: 5.503 ug/l

RT: 11.604 min Scan# 11

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion:112 Resp: 28724

Ion Ratio Lower Upper

112 100

114 32.9 24.6 37.0



#66

1,1,1,2-Tetrachloroethane

Concen: 5.426 ug/l

RT: 11.681 min Scan# 1716

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion:131 Resp: 10175


Ion Ratio Lower Upper

131 100

133 102.9 48.8 146.4

119 74.7 34.8 104.3









#67
Ethyl Benzene
Concen: 4.912 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 43370
Ion Ratio Lower Upper
91 100
106 30.3 23.6 35.4



#68
m/p-Xylenes
Concen: 9.542 ug/l
RT: 11.793 min Scan# 1735
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:106 Resp: 32330
Ion Ratio Lower Upper
106 100
91 207.3 162.6 244.0









#69
o-Xylene
Concen: 4.880 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:106 Resp: 15299
Ion Ratio Lower Upper
106 100
91 211.8 110.5 331.5



#70
Styrene
Concen: 4.697 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:104 Resp: 26200
Ion Ratio Lower Upper
104 100
78 46.8 39.8 59.8
103 55.7 45.2 67.8





#71

Bromoform

Concen: 5.725 ug/l

RT: 12.298 min Scan# 1820

Delta R.T. 0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:173 Resp: 6457

Ion Ratio Lower Upper

173 100

175 47.2 24.8 74.3

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.516 min Scan# 2028

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion:152 Resp: 115588


Ion Ratio Lower Upper

152 100

115 59.7 28.9 86.8

150 158.7 0.0 354.8





#73

Isopropylbenzene

Concen: 5.127 ug/l

RT: 12.422 min Scan# 1809

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion:105 Resp: 40318

Ion Ratio Lower Upper

105 100

120 25.3 13.3 39.9



#74

N-ethyl acetate

Concen: 5.203 ug/l

RT: 12.228 min Scan# 1809

Delta R.T. -0.006 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Tgt Ion: 43 Resp: 9895




Ion Ratio Lower Upper


43 100

70 43.4 36.3 54.5

55 26.7 20.3 30.5

61 22.9 19.7 29.5









#75
1,1,2,2-Tetrachloroethane
Concen: 5.762 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 9187
Ion Ratio Lower Upper
83 100
131 12.7 5.1 15.3
85 63.6 32.9 98.6



#76
1,2,3-Trichloropropane
Concen: 9.650 ug/l
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 75 Resp: 11840
Ion Ratio Lower Upper
75 100
77 167.9 0.0 0.0#





#77

Bromobenzene

Concen: 5.710 ug/l

RT: 12.698 min Scan# 1889

Delta R.T. -0.000 min

Lab File: VD079961.D





Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :

Tgt Ion:	156	Resp:	11615
Ion	Ratio	Lower	Upper
156	100		
77	171.1	89.8	269.6
158	91.2	47.9	143.8



#78

n-propylbenzene

Concen: 5.052 ug/l




RT: 12.763 min Scan# 1900


Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Tgt Ion:	91	Resp:	49389
Ion	Ratio	Lower	Upper
91	100		
120	23.4	11.2	33.5





#79

2-Chlorotoluene

Concen: 5.196 ug/l

RT: 12.845 min Scan# 1914

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D





ClientSampleId :

Tgt Ion: 91 Resp: 29511

Ion Ratio Lower Upper

91 100

126 36.5 17.2 51.5



#80

1,3,5-Trimethylbenzene

Concen: 5.030 ug/l

RT: 12.904 min Scan# 1924

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29


Tgt Ion: 105 Resp: 33333

Ion Ratio Lower Upper

105 100

120 49.3 25.0 75.0









#81
trans-1,4-Dichloro-2-butene
Concen: 5.392 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 3081
Ion Ratio Lower Upper
75 100
53 94.0 68.7 103.1
89 52.2 36.1 54.1



#82
4-Chlorotoluene
Concen: 5.160 ug/l
RT: 12.951 min Scan# 1932
Delta R.T. 0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 91 Resp: 31487
Ion Ratio Lower Upper
91 100
126 36.6 16.8 50.3





#83

tert-Butylbenzene

Concen: 5.015 ug/l

RT: 13.163 min Scan# 1968

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:119 Resp: 28485

Ion Ratio Lower Upper

119 100

91 64.7 33.3 99.9

134 29.7 13.5 40.5



#84

1,2,4-Trimethylbenzene

Concen: 4.781 ug/l

RT: 13.210 min Scan# 1976

Delta R.T. -0.000 min

Lab File: VD079961.D




Acq: 11 Oct 2024 10:29


Tgt Ion:105 Resp: 31814

Ion Ratio Lower Upper

105 100

120 48.7 22.8 68.4









#85
 sec-Butylbenzene
 Concen: 4.838 ug/l
 RT: 13.340 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:105 Resp: 42231
 Ion Ratio Lower Upper
 105 100
 134 20.6 9.8 29.3



#86
 p-Isopropyltoluene
 Concen: 4.842 ug/l
 RT: 13.457 min Scan# 2018
 Delta R.T. -0.000 min
 Lab File: VD079961.D
 Acq: 11 Oct 2024 10:29

Tgt Ion:119 Resp: 35260
 Ion Ratio Lower Upper
 119 100
 134 26.4 13.1 39.1
 91 24.0 12.2 36.4





#87

1,3-Dichlorobenzene

Concen: 5.499 ug/l

RT: 13.457 min Scan# 2018

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Instrument :

MSVOA_D

ClientSampleId :





Tgt Ion:146 Resp: 22886

Ion Ratio Lower Upper

146 100

111 39.6 20.0 59.9

148 65.3 31.9 95.7



#88

1,4-Dichlorobenzene

Concen: 5.796 ug/l

RT: 13.534 min Scan# 2031

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Tgt Ion:146 Resp: 24101


Ion Ratio Lower Upper

146 100

111 43.3 19.3 57.8

148 64.7 32.9 98.7









#89
n-Butylbenzene
Concen: 5.214 ug/l
RT: 13.787 min Scan# 2074
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 36729
Ion Ratio Lower Upper
91 100
92 47.9 26.3 78.9
134 24.5 12.3 36.8



#90
Hexachloroethane
Concen: 5.787 ug/l
RT: 14.051 min Scan# 2119
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29


Tgt Ion: 117 Resp: 9316
Ion Ratio Lower Upper
117 100
201 76.1 37.5 112.5








#91
1,2-Dichlorobenzene
Concen: 5.482 ug/l
RT: 13.828 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:146 Resp: 19930
Ion Ratio Lower Upper
146 100
111 40.3 20.6 61.8
148 65.9 32.7 98.1



#92
1,2-Dibromo-3-Chloropropane
Concen: 6.089 ug/l
RT: 14.434 min Scan# 2184
Delta R.T. -0.006 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion: 75 Resp: 1621
Ion Ratio Lower Upper
75 100
155 89.6 40.6 122.0
157 92.1 53.3 159.9









#93
1,2,4-Trichlorobenzene
Concen: 5.208 ug/l
RT: 15.098 min Scan# 211
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:180 Resp: 11867
Ion Ratio Lower Upper
180 100
182 96.1 47.4 142.2
145 29.4 14.7 44.1



#94
Hexachlorobutadiene
Concen: 5.299 ug/l
RT: 15.198 min Scan# 2314
Delta R.T. -0.000 min
Lab File: VD079961.D
Acq: 11 Oct 2024 10:29

Tgt Ion:225 Resp: 6649
Ion Ratio Lower Upper
225 100
223 68.2 31.5 94.5
227 63.7 30.9 92.5





#95

Naphthalene

Concen: 5.002 ug/l

RT: 15.328 min Scan# 2336

Delta R.T. -0.000 min


Lab File: VD079961.D

Acq: 11 Oct 2024 10:29




Instrument :

MSVOA_D

ClientSampleId :



Tgt Ion:	128	Resp:	20324
Ion Ratio	Lower	Upper	
128	100		
127	13.5	11.8	17.8
129	14.8	8.3	12.5#



#96

1,2,3-Trichlorobenzene

Concen: 5.572 ug/l

RT: 15.516 min Scan# 2368

Delta R.T. -0.000 min

Lab File: VD079961.D

Acq: 11 Oct 2024 10:29

Tgt Ion:	180	Resp:	11252
Ion Ratio	Lower	Upper	
180	100		
182	85.9	47.5	142.6
145	31.5	15.7	47.0

