

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101124\
 Data File : VD079967.D
 Acq On : 11 Oct 2024 16:05
 Operator : RP/MD
 Sample : VIBLK
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Oct 14 03:22:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 14 03:20:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	166613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	278913	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	243280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	109070	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	83915	44.276	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.560%
35) Dibromofluoromethane	7.799	113	90041	44.727	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.460%
50) Toluene-d8	10.269	98	308750	43.039	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.080%
62) 4-Bromofluorobenzene	12.569	95	94265	37.642	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101124\
Data File : VD079967.D
Acq On : 11 Oct 2024 16:05
Operator : RP/MD
Sample : VIBLK
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Oct 14 03:22:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101124S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 14 03:20:54 2024
Response via : Initial Calibration

