

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
 Data File : VD074550.D
 Acq On : 17 Oct 2022 17:47
 Operator : VA/SY
 Sample : N5140-07
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-04-3.5-4.0

Quant Time: Oct 17 23:09:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	3550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	6997	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	6155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	3328	108.912	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 217.820%#	
35) Dibromofluoromethane	7.810	113	2588	62.068	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.140%	
50) Toluene-d8	10.263	98	7966	49.741	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.480%	
62) 4-Bromofluorobenzene	12.569	95	2627	53.813	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 107.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
Data File : VD074550.D
Acq On : 17 Oct 2022 17:47
Operator : VA/SY
Sample : N5140-07
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-04-3.5-4.0

Quant Time: Oct 17 23:09:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

