

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
 Data File : VD074552.D
 Acq On : 17 Oct 2022 18:43
 Operator : VA/SY
 Sample : N5140-11
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-3.5-4.0

Quant Time: Oct 17 23:10:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	100157	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	177905	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	155587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	61097	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	48276	55.998	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.000%
35) Dibromofluoromethane	7.804	113	58866	55.219	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.440%
50) Toluene-d8	10.269	98	226594	55.443	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.880%
62) 4-Bromofluorobenzene	12.569	95	58657	46.508	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
Data File : VD074552.D
Acq On : 17 Oct 2022 18:43
Operator : VA/SY
Sample : N5140-11
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-3.5-4.0

Quant Time: Oct 17 23:10:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

