

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
 Data File : VD074546.D
 Acq On : 17 Oct 2022 15:55
 Operator : VA/SY
 Sample : N5140-03
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-02-3.5-4.0

Quant Time: Oct 17 23:09:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	117374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	193696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	168351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	72244	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	48692	48.195	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.400%
35) Dibromofluoromethane	7.810	113	49329	41.829	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.660%
50) Toluene-d8	10.269	98	243890	54.829	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.660%
62) 4-Bromofluorobenzene	12.569	95	65966	48.241	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101722\
Data File : VD074546.D
Acq On : 17 Oct 2022 15:55
Operator : VA/SY
Sample : N5140-03
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-02-3.5-4.0

Quant Time: Oct 17 23:09:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55




Instrument :
MSVOA_D
ClientSampleId :
SB-02-3.5-4.0


Tgt Ion:168 Resp: 117374
Ion Ratio Lower Upper
168 100
99 51.5 40.9 61.3



#33
1,2-Dichloroethane-d4
Concen: 48.195 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55

Tgt Ion: 65 Resp: 48692
Ion Ratio Lower Upper
65 100
67 55.4 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074546.D

Acq: 17 Oct 2022 15:55

Instrument :

MSVOA_D

ClientSampleId :

SB-02-3.5-4.0





Tgt Ion:114 Resp: 193696

Ion Ratio Lower Upper

114 100

63 15.9 0.0 33.2

88 13.4 0.0 28.2



#35

Dibromofluoromethane

Concen: 41.829 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD074546.D

Acq: 17 Oct 2022 15:55




Tgt Ion:113 Resp: 49329


Ion Ratio Lower Upper

113 100

111 105.2 82.3 123.5

192 19.3 16.4 24.6









#50
Toluene-d8
Concen: 54.829 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55




Instrument :
MSVOA_D
ClientSampleId :
SB-02-3.5-4.0


Tgt Ion: 98 Resp: 243890
Ion Ratio Lower Upper
98 100
100 63.2 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 48.241 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55

Tgt Ion: 95 Resp: 65966
Ion Ratio Lower Upper
95 100
174 94.2 0.0 197.0
176 90.9 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55

Instrument :
MSVOA_D
ClientSampleId :
SB-02-3.5-4.0

Tgt Ion:117 Resp: 168351
Ion Ratio Lower Upper
117 100
82 52.0 41.6 62.4
119 32.4 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.006 min
Lab File: VD074546.D
Acq: 17 Oct 2022 15:55

Tgt Ion:152 Resp: 72244
Ion Ratio Lower Upper
152 100
115 55.1 26.8 80.4
150 154.5 0.0 348.2

