


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD101819\
 Data File : VD064001.D
 Acq On : 18 Oct 2019 10:18
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100219S.M
 Title : SW846 8260
 Last Update : Sat Oct 05 03:51:17 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	24216	PASS
75	95	30	60	54.4	62154	PASS
95	95	100	100	100.0	114173	PASS
96	95	5	9	6.6	7525	PASS
173	174	0.00	2	1.9	2117	PASS
174	95	50	100	95.4	108918	PASS
175	174	5	9	8.2	8954	PASS
176	174	95	101	97.4	106122	PASS
177	176	5	9	7.4	7893	PASS