

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070860.D
 Acq On : 27 Oct 2021 13:26
 Operator : VA/SY
 Sample : M4356-01RE
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-282RE

Quant Time: Oct 28 03:55:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	93387	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	164908	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	177885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	77157	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	80605	85.819	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	171.640%#
35) Dibromofluoromethane	7.908	113	73744	67.822	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	135.640%
50) Toluene-d8	10.332	98	213829	52.272	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.540%
62) 4-Bromofluorobenzene	12.620	95	69090	51.189	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
Data File : VD070860.D
Acq On : 27 Oct 2021 13:26
Operator : VA/SY
Sample : M4356-01RE
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-282RE

Quant Time: Oct 28 03:55:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

