

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070865.D
 Acq On : 27 Oct 2021 15:48
 Operator : VA/SY
 Sample : M4368-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-02-102621

Quant Time: Oct 28 03:56:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	24195	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	40134	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	43898	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	19014	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	22071	90.700	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	181.400%#
35) Dibromofluoromethane	7.909	113	16199	61.215	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.440%
50) Toluene-d8	10.332	98	50573	50.799	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.600%
62) 4-Bromofluorobenzene	12.620	95	17649	53.729	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
Data File : VD070865.D
Acq On : 27 Oct 2021 15:48
Operator : VA/SY
Sample : M4368-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-02-102621

Quant Time: Oct 28 03:56:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

