

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070870.D
 Acq On : 27 Oct 2021 18:09
 Operator : VA/SY
 Sample : M4372-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-1

Quant Time: Oct 28 03:58:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	414494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	741867	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	702576	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	287685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	210551	50.507	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.020%
35) Dibromofluoromethane	7.909	113	251781	51.473	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.940%
50) Toluene-d8	10.332	98	947620	51.494	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.980%
62) 4-Bromofluorobenzene	12.620	95	266084	43.822	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.640%
Target Compounds						
80) 1,3,5-Trimethylbenzene	12.950	105	18826	1.156	ug/l	94
84) 1,2,4-Trimethylbenzene	13.255	105	39186	2.440	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
Data File : VD070870.D
Acq On : 27 Oct 2021 18:09
Operator : VA/SY
Sample : M4372-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-1

Quant Time: Oct 28 03:58:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

