

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070878.D
 Acq On : 27 Oct 2021 23:20
 Operator : VA/SY
 Sample : VD1027SBL02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1027SBL02

Quant Time: Oct 28 04:03:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.979	168	427830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	747016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	696257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	332138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	205706	47.806	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.620%
35) Dibromofluoromethane	7.908	113	255157	51.804	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.600%
50) Toluene-d8	10.332	98	908724	49.040	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.620	95	291734	47.716	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070878.D
 Acq On : 27 Oct 2021 23:20
 Operator : VA/SY
 Sample : VD1027SBL02
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1027SBL02

Quant Time: Oct 28 04:03:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

