

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070877.D
 Acq On : 27 Oct 2021 22:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 28 04:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	120	0.00
2 T	Dichlorodifluoromethane	0.464	0.383	17.5	106	0.00
3 P	Chloromethane	0.656	0.563	14.2	103	0.00
4 C	Vinyl Chloride	0.726	0.640	11.8#	107	0.00
5 T	Bromomethane	0.520	0.401	22.9	100	0.00
6 T	Chloroethane	0.449	0.344	23.4	94	0.00
7 T	Trichlorofluoromethane	0.894	0.756	15.4	104	0.00
8 T	Diethyl Ether	0.248	0.256	-3.2	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.576	0.491	14.8	108	0.00
10 T	Methyl Iodide	0.511	0.631	-23.5	118	0.00
11 T	Tert butyl alcohol	0.027	0.025	7.4	123	0.00
12 CM	1,1-Dichloroethene	0.505	0.484	4.2#	111	0.00
13 T	Acrolein	0.024	0.021	12.5	105	0.00
14 T	Allyl chloride	0.790	0.763	3.4	106	0.00
15 T	Acrylonitrile	0.119	0.112	5.9	108	0.00
16 T	Acetone	0.117	0.082	29.9#	74	0.00
17 T	Carbon Disulfide	1.784	1.659	7.0	107	0.00
18 T	Methyl Acetate	0.401	0.393	2.0	118	0.00
19 T	Methyl tert-butyl Ether	1.049	1.118	-6.6	119	0.00
20 T	Methylene Chloride	0.809	0.715	11.6	129	0.00
21 T	trans-1,2-Dichloroethene	0.598	0.596	0.3	114	0.00
22 T	Diisopropyl ether	1.639	1.655	-1.0	109	0.00
23 T	Vinyl Acetate	0.745	0.738	0.9	106	0.00
24 P	1,1-Dichloroethane	1.105	1.043	5.6	112	0.00
25 T	2-Butanone	0.144	0.125	13.2	96	0.00
26 T	2,2-Dichloropropane	0.881	0.762	13.5	101	0.00
27 T	cis-1,2-Dichloroethene	0.641	0.658	-2.7	117	0.00
28 T	Bromochloromethane	0.433	0.416	3.9	111	0.00
29 T	Tetrahydrofuran	0.087	0.087	0.0	111	0.00
30 C	Chloroform	1.109	1.052	5.1#	114	0.00
31 T	Cyclohexane	1.025	0.895	12.7	102	0.00
32 T	1,1,1-Trichloroethane	0.947	0.846	10.7	106	0.00
33 S	1,2-Dichloroethane-d4	0.503	0.485	3.6	121	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	127	0.00
35 S	Dibromofluoromethane	0.330	0.329	0.3	127	0.00
36 T	1,1-Dichloropropene	0.523	0.482	7.8	109	0.00
37 T	Ethyl Acetate	0.195	0.173	11.3	102	0.00
38 T	Carbon Tetrachloride	0.526	0.457	13.1	107	0.00
39 T	Methylcyclohexane	0.602	0.561	6.8	107	0.00
40 TM	Benzene	1.521	1.435	5.7	111	0.00
41 T	Methacrylonitrile	0.103	0.111	-7.8	145	0.00
42 TM	1,2-Dichloroethane	0.387	0.351	9.3	113	0.00
43 T	Isopropyl Acetate	0.362	0.333	8.0	110	0.00
44 TM	Trichloroethene	0.393	0.382	2.8	121	0.00
45 C	1,2-Dichloropropane	0.403	0.379	6.0#	114	0.00
46 T	Dibromomethane	0.198	0.189	4.5	116	0.00
47 T	Bromodichloromethane	0.511	0.466	8.8	113	0.00
48 T	Methyl methacrylate	0.173	0.167	3.5	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070877.D
 Acq On : 27 Oct 2021 22:23
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 28 04:02:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 27 05:34:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	123	0.00
50 S	Toluene-d8	1.240	1.260	-1.6	125	0.00
51 T	4-Methyl-2-Pentanone	0.190	0.180	5.3	109	0.00
52 CM	Toluene	0.914	0.903	1.2#	113	0.00
53 T	t-1,3-Dichloropropene	0.445	0.430	3.4	114	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.526	4.4	117	0.00
55 T	1,1,2-Trichloroethane	0.280	0.271	3.2	120	0.00
56 T	Ethyl methacrylate	0.308	0.324	-5.2	119	0.00
57 T	1,3-Dichloropropane	0.472	0.462	2.1	117	0.00
58 T	2-Chloroethyl Vinyl ether	0.133	0.135	-1.5	127	0.00
59 T	2-Hexanone	0.135	0.121	10.4	99	0.00
60 T	Dibromochloromethane	0.336	0.329	2.1	120	0.00
61 T	1,2-Dibromoethane	0.259	0.253	2.3	118	0.00
62 S	4-Bromofluorobenzene	0.409	0.420	-2.7	128	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	125	0.00
64 T	Tetrachloroethene	0.343	0.351	-2.3	119	0.00
65 PM	Chlorobenzene	1.026	1.029	-0.3	118	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.377	1.0	117	0.00
67 C	Ethyl Benzene	1.786	1.835	-2.7#	115	0.00
68 T	m/p-Xylenes	0.696	0.724	-4.0	115	0.00
69 T	o-Xylene	0.630	0.667	-5.9	117	0.00
70 T	Styrene	1.091	1.154	-5.8	115	0.00
71 P	Bromoform	0.201	0.206	-2.5	120	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
73 T	Isopropylbenzene	3.335	3.548	-6.4	114	0.00
74 T	N-amyl acetate	0.733	0.696	5.0	105	0.00
75 P	1,1,2,2-Tetrachloroethane	0.681	0.663	2.6	115	0.00
76 T	1,2,3-Trichloropropane	0.492	0.498	-1.2	115	0.00
77 T	Bromobenzene	0.813	0.849	-4.4	119	0.00
78 T	n-propylbenzene	4.281	4.463	-4.3	113	0.00
79 T	2-Chlorotoluene	2.486	2.557	-2.9	114	0.00
80 T	1,3,5-Trimethylbenzene	2.830	3.024	-6.9	114	0.00
81 T	trans-1,4-Dichloro-2-butene	0.188	0.185	1.6	109	0.00
82 T	4-Chlorotoluene	2.583	2.689	-4.1	115	0.00
83 T	tert-Butylbenzene	2.360	2.517	-6.7	116	0.00
84 T	1,2,4-Trimethylbenzene	2.791	2.991	-7.2	114	0.00
85 T	sec-Butylbenzene	3.723	3.804	-2.2	112	0.00
86 T	p-Isopropyltoluene	2.995	3.145	-5.0	113	0.00
87 T	1,3-Dichlorobenzene	1.678	1.693	-0.9	116	0.00
88 T	1,4-Dichlorobenzene	1.664	1.653	0.7	115	0.00
89 T	n-Butylbenzene	2.934	2.930	0.1	111	0.00
90 T	Hexachloroethane	0.672	0.625	7.0	113	0.00
91 T	1,2-Dichlorobenzene	1.432	1.462	-2.1	117	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.099	0.090	9.1	107	0.00
93 T	1,2,4-Trichlorobenzene	0.827	0.879	-6.3	123	0.00
94 T	Hexachlorobutadiene	0.545	0.540	0.9	119	0.00
95 T	Naphthalene	1.393	1.546	-11.0	123	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
Data File : VD070877.D
Acq On : 27 Oct 2021 22:23
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 28 04:02:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 27 05:34:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.719	0.767	-6.7	119	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6