



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102721\
 Data File : VD070852.D
 Acq On : 27 Oct 2021 09:24
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102621S.M
 Title : SW846 8260
 Last Update : Wed Oct 27 05:34:47 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	26096	PASS
75	95	30	60	46.6	72685	PASS
95	95	100	100	100.0	155848	PASS
96	95	5	9	6.1	9530	PASS
173	174	0.00	2	1.2	1549	PASS
174	95	50	100	83.2	129680	PASS
175	174	5	9	8.1	10560	PASS
176	174	95	101	98.3	127453	PASS
177	176	5	9	6.3	8083	PASS