

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102722\
 Data File : VD074660.D
 Acq On : 27 Oct 2022 19:14
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Oct 28 00:22:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 00:18:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	42329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	148776	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	168741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	85464	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	37071	103.763	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	207.520%#
35) Dibromofluoromethane	7.805	113	58374	64.256	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.520%
50) Toluene-d8	10.269	98	182503	52.238	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.480%
62) 4-Bromofluorobenzene	12.569	95	68635	71.533	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	143.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102722\
Data File : VD074660.D
Acq On : 27 Oct 2022 19:14
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Oct 28 00:22:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D102722S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 00:18:17 2022
Response via : Initial Calibration

