

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
 Data File : VD074714.D
 Acq On : 03 Nov 2022 14:41
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Nov 04 00:02:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	142426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	244018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	191328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	81796	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	59799	53.011	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.020%
35) Dibromofluoromethane	7.804	113	74927	51.908	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.820%
50) Toluene-d8	10.269	98	266659	50.644	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.280%
62) 4-Bromofluorobenzene	12.575	95	70066	46.088	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
Data File : VD074714.D
Acq On : 03 Nov 2022 14:41
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Nov 04 00:02:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

