

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
 Data File : VD074719.D
 Acq On : 03 Nov 2022 17:00
 Operator : VA/SY
 Sample : N5396-16
 Misc : 4.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-H

Quant Time: Nov 04 00:03:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	163337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	283789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	221749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	99579	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	76595	59.208	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.420%
35) Dibromofluoromethane	7.805	113	96622	57.557	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.120%
50) Toluene-d8	10.269	98	303843	49.619	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.569	95	89525	50.635	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
Data File : VD074719.D
Acq On : 03 Nov 2022 17:00
Operator : VA/SY
Sample : N5396-16
Misc : 4.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-H

Quant Time: Nov 04 00:03:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

