

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
 Data File : VD074721.D
 Acq On : 03 Nov 2022 17:55
 Operator : VA/SY
 Sample : N5430-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SOIL-COMP-1102

Quant Time: Nov 04 00:03:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	9744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	18669	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	19221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	9325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	11237	145.605	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 291.220%#	
35) Dibromofluoromethane	7.805	113	9767	88.441	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 176.880%#	
50) Toluene-d8	10.269	98	23036	57.185	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 114.360%	
62) 4-Bromofluorobenzene	12.569	95	8973	77.147	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 154.300%#	
Target Compounds						
16) Acetone	4.022	43	2094	130.022	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.569	75	4102	134.699	ug/l #	17

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
Data File : VD074721.D
Acq On : 03 Nov 2022 17:55
Operator : VA/SY
Sample : N5430-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SOIL-COMP-1102

Quant Time: Nov 04 00:03:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

