

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
 Data File : VD074726.D
 Acq On : 03 Nov 2022 20:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 00:04:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.411	0.367	10.7	94	0.00
3 P	Chloromethane	0.655	0.606	7.5	94	0.00
4 C	Vinyl Chloride	0.719	0.704	2.1#	95	0.00
5 T	Bromomethane	0.545	0.530	2.8	105	0.00
6 T	Chloroethane	0.498	0.499	-0.2	98	0.00
7 T	Trichlorofluoromethane	0.835	0.810	3.0	94	0.00
8 T	Diethyl Ether	0.244	0.247	-1.2	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.530	0.532	-0.4	99	0.00
10 T	Methyl Iodide	0.571	0.624	-9.3	93	0.00
11 T	Tert butyl alcohol	0.070	0.034	51.4#	89	0.00
12 CM	1,1-Dichloroethene	0.539	0.528	2.0#	92	0.00
13 T	Acrolein	0.027	0.021	22.2	80	0.00
14 T	Allyl chloride	0.570	0.592	-3.9	99	0.00
15 T	Acrylonitrile	0.096	0.095	1.0	94	0.00
16 T	Acetone	0.083	0.069	16.9	79	0.00
17 T	Carbon Disulfide	1.624	1.624	0.0	94	0.00
18 T	Methyl Acetate	0.250	0.225	10.0	96	0.00
19 T	Methyl tert-butyl Ether	1.024	1.086	-6.1	100	0.00
20 T	Methylene Chloride	1.061	0.731	31.1#	102	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.623	-5.2	100	0.00
22 T	Diisopropyl ether	1.176	1.284	-9.2	100	0.00
23 T	Vinyl Acetate	0.500	0.528	-5.6	101	0.00
24 P	1,1-Dichloroethane	0.900	0.929	-3.2	100	0.00
25 T	2-Butanone	0.114	0.104	8.8	89	0.00
26 T	2,2-Dichloropropane	0.828	0.796	3.9	95	0.00
27 T	cis-1,2-Dichloroethene	0.653	0.670	-2.6	98	0.00
28 T	Bromochloromethane	0.284	0.302	-6.3	104	0.00
29 T	Tetrahydrofuran	0.065	0.067	-3.1	96	0.00
30 C	Chloroform	0.961	1.012	-5.3#	102	0.00
31 T	Cyclohexane	0.802	0.770	4.0	94	0.00
32 T	1,1,1-Trichloroethane	0.849	0.882	-3.9	99	0.00
33 S	1,2-Dichloroethane-d4	0.396	0.377	4.8	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.296	0.285	3.7	101	0.00
36 T	1,1-Dichloropropene	0.449	0.449	0.0	96	0.00
37 T	Ethyl Acetate	0.150	0.152	-1.3	100	0.00
38 T	Carbon Tetrachloride	0.425	0.430	-1.2	98	0.00
39 T	Methylcyclohexane	0.546	0.519	4.9	88	0.00
40 TM	Benzene	1.327	1.351	-1.8	98	0.00
41 T	Methacrylonitrile	0.079	0.081	-2.5	94	0.00
42 TM	1,2-Dichloroethane	0.305	0.314	-3.0	100	0.00
43 T	Isopropyl Acetate	0.279	0.282	-1.1	99	0.00
44 TM	Trichloroethene	0.378	0.383	-1.3	98	0.00
45 C	1,2-Dichloropropane	0.293	0.284	3.1#	91	0.00
46 T	Dibromomethane	0.170	0.161	5.3	90	0.00
47 T	Bromodichloromethane	0.409	0.398	2.7	91	0.00
48 T	Methyl methacrylate	0.124	0.120	3.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
 Data File : VD074726.D
 Acq On : 03 Nov 2022 20:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 04 00:04:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	1.079	0.955	11.5	98	0.00
51 T	4-Methyl-2-Pentanone	0.135	0.130	3.7	91	0.00
52 CM	Toluene	0.812	0.786	3.2#	98	0.00
53 T	t-1,3-Dichloropropene	0.362	0.353	2.5	99	0.00
54 T	cis-1,3-Dichloropropene	0.454	0.439	3.3	89	0.00
55 T	1,1,2-Trichloroethane	0.231	0.221	4.3	100	0.00
56 T	Ethyl methacrylate	0.258	0.256	0.8	100	0.00
57 T	1,3-Dichloropropane	0.372	0.359	3.5	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.068	0.068	0.0	95	0.00
59 T	2-Hexanone	0.093	0.087	6.5	93	0.00
60 T	Dibromochloromethane	0.276	0.271	1.8	104	0.00
61 T	1,2-Dibromoethane	0.218	0.210	3.7	101	0.00
62 S	4-Bromofluorobenzene	0.312	0.285	8.7	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.353	0.351	0.6	101	0.00
65 PM	Chlorobenzene	0.990	1.040	-5.1	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.336	0.363	-8.0	103	0.00
67 C	Ethyl Benzene	1.755	1.857	-5.8#	99	0.00
68 T	m/p-Xylenes	0.697	0.739	-6.0	99	0.00
69 T	o-Xylene	0.640	0.695	-8.6	100	0.00
70 T	Styrene	1.085	1.187	-9.4	101	0.00
71 P	Bromoform	0.180	0.190	-5.6	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.653	3.855	-5.5	99	0.00
74 T	N-amyl acetate	0.593	0.615	-3.7	98	0.00
75 P	1,1,2,2-Tetrachloroethane	0.614	0.642	-4.6	102	0.00
76 T	1,2,3-Trichloropropane	0.388	0.388	0.0	96	0.00
77 T	Bromobenzene	0.841	0.904	-7.5	104	0.00
78 T	n-propylbenzene	4.376	4.631	-5.8	99	0.00
79 T	2-Chlorotoluene	2.356	2.505	-6.3	101	0.00
80 T	1,3,5-Trimethylbenzene	2.980	3.171	-6.4	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.163	0.173	-6.1	100	0.00
82 T	4-Chlorotoluene	2.451	2.589	-5.6	101	0.00
83 T	tert-Butylbenzene	2.626	2.687	-2.3	95	0.00
84 T	1,2,4-Trimethylbenzene	2.941	3.165	-7.6	101	0.00
85 T	sec-Butylbenzene	3.962	4.151	-4.8	99	0.00
86 T	p-Isopropyltoluene	3.337	3.497	-4.8	99	0.00
87 T	1,3-Dichlorobenzene	1.744	1.809	-3.7	101	0.00
88 T	1,4-Dichlorobenzene	1.715	1.741	-1.5	99	0.00
89 T	n-Butylbenzene	3.052	3.188	-4.5	98	0.00
90 T	Hexachloroethane	0.572	0.597	-4.4	98	0.00
91 T	1,2-Dichlorobenzene	1.486	1.561	-5.0	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.084	1.2	99	0.00
93 T	1,2,4-Trichlorobenzene	0.917	0.941	-2.6	100	0.00
94 T	Hexachlorobutadiene	0.480	0.492	-2.5	99	0.00
95 T	Naphthalene	1.762	1.814	-3.0	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
Data File : VD074726.D
Acq On : 03 Nov 2022 20:15
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 04 00:04:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.807	0.839	-4.0	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6