



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110322\
Data File : VD074707.D
Acq On : 03 Nov 2022 09:01
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Title : SW846 8260
Last Update : Thu Nov 03 02:30:45 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	10500	PASS
75	95	30	60	49.0	33592	PASS
95	95	100	100	100.0	68488	PASS
96	95	5	9	6.8	4664	PASS
173	174	0.00	2	0.9	568	PASS
174	95	50	100	92.6	63408	PASS
175	174	5	9	7.6	4824	PASS
176	174	95	101	97.2	61664	PASS
177	176	5	9	6.9	4234	PASS