

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
 Data File : VD074800.D
 Acq On : 08 Nov 2022 13:17
 Operator : VA/SY
 Sample : N5415-04
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-17

Quant Time: Nov 08 14:49:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	44717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	76625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	81087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	36246	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	35856	101.240	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 202.480%#	
35) Dibromofluoromethane	7.805	113	39838	87.891	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 175.780%#	
50) Toluene-d8	10.269	98	98044	59.299	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 118.600%	
62) 4-Bromofluorobenzene	12.569	95	32679	68.454	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 136.900%	
Target Compounds						
16) Acetone	4.017	43	2337	31.620	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
Data File : VD074800.D
Acq On : 08 Nov 2022 13:17
Operator : VA/SY
Sample : N5415-04
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-17

Quant Time: Nov 08 14:49:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

