

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
 Data File : VD074801.D
 Acq On : 08 Nov 2022 13:44
 Operator : VA/SY
 Sample : N5415-05
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-18

Quant Time: Nov 08 14:49:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	20816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	35627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	38137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	19108	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	20002	121.322	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	242.640%#
35) Dibromofluoromethane	7.805	113	20108	95.413	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	190.820%#
50) Toluene-d8	10.269	98	44084	57.345	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.700%
62) 4-Bromofluorobenzene	12.569	95	16737	75.405	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	150.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
Data File : VD074801.D
Acq On : 08 Nov 2022 13:44
Operator : VA/SY
Sample : N5415-05
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-18

Quant Time: Nov 08 14:49:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

