

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
 Data File : VD074805.D
 Acq On : 08 Nov 2022 15:35
 Operator : VA/SY
 Sample : N5415-11RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R-24RE

Quant Time: Nov 08 16:31:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	32268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	60006	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	60917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	28150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	31365	122.726	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	245.460%#	
35) Dibromofluoromethane	7.810	113	29310	82.573	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	165.140%#	
50) Toluene-d8	10.269	98	67731	52.310	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	104.620%	
62) 4-Bromofluorobenzene	12.569	95	23956	64.080	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	128.160%	
Target Compounds						Qvalue
20) Methylene Chloride	4.799	84	7897	8.034	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
Data File : VD074805.D
Acq On : 08 Nov 2022 15:35
Operator : VA/SY
Sample : N5415-11RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R-24RE

Quant Time: Nov 08 16:31:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

