

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
 Data File : VD074810.D
 Acq On : 08 Nov 2022 18:21
 Operator : VA/SY
 Sample : N5494-07RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-16RE

Quant Time: Nov 08 23:39:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	1264	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	952	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	335	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	534	89.065	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 178.140%	#
35) Dibromofluoromethane	7.805	113	498	66.604	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 133.200%	
50) Toluene-d8	10.275	98	1276	46.784	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.560%	
62) 4-Bromofluorobenzene	12.575	95	205	26.032	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 52.060%	
Target Compounds						Qvalue
76) 1,2,3-Trichloropropane	12.575	75	67	25.800	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.575	75	67	61.242	ug/l #	17

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
Data File : VD074810.D
Acq On : 08 Nov 2022 18:21
Operator : VA/SY
Sample : N5494-07RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-16RE

Quant Time: Nov 08 23:39:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

