

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
 Data File : VD074811.D
 Acq On : 08 Nov 2022 18:48
 Operator : VA/SY
 Sample : N5494-11
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-24

Quant Time: Nov 08 23:39:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	128468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	194295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	178329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	74895	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	55994	55.031	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.060%
35) Dibromofluoromethane	7.805	113	80389	69.944	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.880%
50) Toluene-d8	10.269	98	236840	56.492	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.980%
62) 4-Bromofluorobenzene	12.575	95	67389	55.671	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110822\
Data File : VD074811.D
Acq On : 08 Nov 2022 18:48
Operator : VA/SY
Sample : N5494-11
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-24

Quant Time: Nov 08 23:39:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

