

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074823.D
 Acq On : 09 Nov 2022 12:59
 Operator : VA/SY
 Sample : N5485-01RE
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-110422RE

Quant Time: Nov 09 14:04:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	182788	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	341919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	299178	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	125786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.233	65	60046	41.476	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.960%
35) Dibromofluoromethane	7.804	113	79009	39.063	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.120%
50) Toluene-d8	10.269	98	261581	35.455	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	70.920%
62) 4-Bromofluorobenzene	12.574	95	79969	37.540	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.080%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074823.D
Acq On : 09 Nov 2022 12:59
Operator : VA/SY
Sample : N5485-01RE
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-110422RE

Quant Time: Nov 09 14:04:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

