

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074827.D
 Acq On : 09 Nov 2022 14:49
 Operator : VA/SY
 Sample : N5505-01
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-2-110722

Quant Time: Nov 10 00:36:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	17310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	34363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	35764	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	15750	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	12384	90.329	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 180.660%#	
35) Dibromofluoromethane	7.810	113	11984	58.956	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.920%	
50) Toluene-d8	10.269	98	29047	39.175	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 78.340%	
62) 4-Bromofluorobenzene	12.575	95	10833	50.601	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.200%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074827.D
Acq On : 09 Nov 2022 14:49
Operator : VA/SY
Sample : N5505-01
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-2-110722

Quant Time: Nov 10 00:36:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

