

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074831.D
 Acq On : 09 Nov 2022 16:38
 Operator : VA/SY
 Sample : N5520-07
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R20-3

Quant Time: Nov 10 00:36:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	161339	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	264731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235240	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	91435	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	59181	46.313	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.620%
35) Dibromofluoromethane	7.811	113	80549	51.436	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.880%
50) Toluene-d8	10.269	98	214334	37.522	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	75.040%
62) 4-Bromofluorobenzene	12.575	95	64160	38.901	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074831.D
Acq On : 09 Nov 2022 16:38
Operator : VA/SY
Sample : N5520-07
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R20-3

Quant Time: Nov 10 00:36:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

