

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074832.D
 Acq On : 09 Nov 2022 17:06
 Operator : VA/SY
 Sample : N5518-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAT-02-110822

Quant Time: Nov 10 00:36:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	1513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	2761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	2566	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	1106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	950	79.277	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 158.560%	
35) Dibromofluoromethane	7.799	113	865	52.962	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.920%	
50) Toluene-d8	10.275	98	2276	38.203	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 76.400%	
62) 4-Bromofluorobenzene	12.575	95	832	48.368	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.740%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074832.D
Acq On : 09 Nov 2022 17:06
Operator : VA/SY
Sample : N5518-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAT-02-110822

Quant Time: Nov 10 00:36:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

