

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074834.D
 Acq On : 09 Nov 2022 18:00
 Operator : VA/SY
 Sample : N5511-03
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-04-110722

Quant Time: Nov 10 00:37:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	190393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	309528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	264443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	98257	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	58852	39.028	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.060%
35) Dibromofluoromethane	7.810	113	87506	47.792	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.580%
50) Toluene-d8	10.269	98	239454	35.852	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	71.700%
62) 4-Bromofluorobenzene	12.575	95	69492	36.036	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	72.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074834.D
Acq On : 09 Nov 2022 18:00
Operator : VA/SY
Sample : N5511-03
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-04-110722

Quant Time: Nov 10 00:37:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

