

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074836.D
 Acq On : 09 Nov 2022 18:55
 Operator : VA/SY
 Sample : N5528-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SAMPLE-3

Quant Time: Nov 10 00:37:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	19335	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	36509	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	31295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	9312	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	15904	103.855	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	207.700%#
35) Dibromofluoromethane	7.810	113	14838	68.705	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	137.420%
50) Toluene-d8	10.269	98	26313	33.402	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	66.800%
62) 4-Bromofluorobenzene	12.575	95	6745	29.654	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	59.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074836.D
Acq On : 09 Nov 2022 18:55
Operator : VA/SY
Sample : N5528-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SAMPLE-3

Quant Time: Nov 10 00:37:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

