

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074848.D
 Acq On : 10 Nov 2022 02:14
 Operator : VA/SY
 Sample : N5506-04
 Misc : 5.55G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-6

Quant Time: Nov 10 05:50:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	4402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	8816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	9071	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	4082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	3665	105.121	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 210.240%#	
35) Dibromofluoromethane	7.810	113	3590	68.840	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 137.680%	
50) Toluene-d8	10.269	98	7943	41.755	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 83.520%	
62) 4-Bromofluorobenzene	12.575	95	2864	52.144	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.280%	
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.798	106	1017	8.047	ug/l	89
84) 1,2,4-Trimethylbenzene	13.210	105	1429	5.952	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074848.D
Acq On : 10 Nov 2022 02:14
Operator : VA/SY
Sample : N5506-04
Misc : 5.55G/5.00ml/MSVOA_D/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-6

Quant Time: Nov 10 05:50:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

