

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074855.D
 Acq On : 10 Nov 2022 05:26
 Operator : VA/SY
 Sample : N5491-10
 Misc : 4.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4

Quant Time: Nov 10 05:51:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	16970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	38754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	41321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	17572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	18826	140.068	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 280.140%#	
35) Dibromofluoromethane	7.805	113	15965	69.642	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 139.280%	
50) Toluene-d8	10.269	98	33841	40.469	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 80.940%	
62) 4-Bromofluorobenzene	12.575	95	12475	51.668	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 103.340%	
Target Compounds						
20) Methylene Chloride	4.805	84	3831	6.300	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074855.D
Acq On : 10 Nov 2022 05:26
Operator : VA/SY
Sample : N5491-10
Misc : 4.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4

Quant Time: Nov 10 05:51:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration

