

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
 Data File : VD074833.D
 Acq On : 09 Nov 2022 17:33
 Operator : VA/SY
 Sample : N5518-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAT-03-110822

Quant Time: Nov 10 00:37:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 03 02:30:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	143647	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	273398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	238357	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	96213	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	58053	51.026	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.060%
35) Dibromofluoromethane	7.810	113	79258	49.008	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.020%
50) Toluene-d8	10.269	98	215913	36.600	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	73.200%
62) 4-Bromofluorobenzene	12.575	95	65141	38.244	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD110922\
Data File : VD074833.D
Acq On : 09 Nov 2022 17:33
Operator : VA/SY
Sample : N5518-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAT-03-110822


Quant Time: Nov 10 00:37:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D110222S.M
Quant Title : SW846 8260
QLast Update : Thu Nov 03 02:30:45 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33


Instrument :
MSVOA_D
ClientSampleId :
PAT-03-110822





Tgt Ion:168 Resp: 143647
Ion Ratio Lower Upper
168 100
99 50.0 40.9 61.3




#33
1,2-Dichloroethane-d4
Concen: 51.026 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.006 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33



Tgt Ion: 65 Resp: 58053
Ion Ratio Lower Upper
65 100
67 54.7 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 11

Delta R.T. 0.006 min

Lab File: VD074833.D

Acq: 09 Nov 2022 17:33

Instrument :

MSVOA_D

ClientSampleId :

PAT-03-110822





Tgt Ion:114 Resp: 273398

Ion Ratio Lower Upper

114 100

63 17.0 0.0 33.2

88 13.3 0.0 28.2



#35

Dibromofluoromethane

Concen: 49.008 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.005 min

Lab File: VD074833.D

Acq: 09 Nov 2022 17:33




Tgt Ion:113 Resp: 79258


Ion Ratio Lower Upper

113 100

111 103.2 82.3 123.5

192 19.0 16.4 24.6









#50
Toluene-d8
Concen: 36.600 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33




Instrument :
MSVOA_D
ClientSampleId :
PAT-03-110822


Tgt Ion: 98 Resp: 215913
Ion Ratio Lower Upper
98 100
100 64.4 51.4 77.0



#62
4-Bromofluorobenzene
Concen: 38.244 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33

Tgt Ion: 95 Resp: 65141
Ion Ratio Lower Upper
95 100
174 95.8 0.0 197.0
176 92.7 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33

Instrument :
MSVOA_D
ClientSampleId :
PAT-03-110822

Tgt Ion:117 Resp: 238357
Ion Ratio Lower Upper
117 100
82 49.6 41.6 62.4
119 32.4 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. 0.000 min
Lab File: VD074833.D
Acq: 09 Nov 2022 17:33

Tgt Ion:152 Resp: 96213
Ion Ratio Lower Upper
152 100
115 54.3 26.8 80.4
150 157.8 0.0 348.2

